

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115536.D **Time Analyzed:** 00:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.475	5.387	5.587	501.710	500.000	0.3
Aroclor-1016-2	5.497	5.408	5.608	480.700	500.000	-3.9
Aroclor-1016-3	5.558	5.470	5.670	530.600	500.000	6.1
Aroclor-1016-4	5.654	5.566	5.766	491.660	500.000	-1.7
Aroclor-1016-5	5.945	5.858	6.058	477.920	500.000	-4.4
Aroclor-1260-1	7.059	6.971	7.171	456.290	500.000	-8.7
Aroclor-1260-2	7.313	7.225	7.425	500.970	500.000	0.2
Aroclor-1260-3	7.671	7.582	7.782	498.060	500.000	-0.4
Aroclor-1260-4	7.894	7.803	8.003	494.120	500.000	-1.2
Aroclor-1260-5	8.201	8.110	8.310	478.800	500.000	-4.2
Decachlorobiphenyl	9.915	9.830	10.030	56.220	50.000	12.4
Tetrachloro-m-xylene	4.337	4.248	4.448	58.070	50.000	16.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115536.D **Time Analyzed:** 00:18

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.628	4.828	480.140	500.000	-4.0
Aroclor-1016-2	4.734	4.647	4.847	467.550	500.000	-6.5
Aroclor-1016-3	4.908	4.821	5.021	472.350	500.000	-5.5
Aroclor-1016-4	4.949	4.864	5.064	303.200	500.000	-39.4
Aroclor-1016-5	5.163	5.076	5.276	452.350	500.000	-9.5
Aroclor-1260-1	6.192	6.106	6.306	421.350	500.000	-15.7
Aroclor-1260-2	6.380	6.294	6.494	407.390	500.000	-18.5
Aroclor-1260-3	6.532	6.446	6.646	412.590	500.000	-17.5
Aroclor-1260-4	7.003	6.917	7.117	405.330	500.000	-18.9
Aroclor-1260-5	7.245	7.158	7.358	406.440	500.000	-18.7
Decachlorobiphenyl	8.625	8.541	8.741	52.940	50.000	5.9
Tetrachloro-m-xylene	3.640	3.548	3.748	54.000	50.000	8.0