

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115552.D **Time Analyzed:** 09:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.474	5.387	5.587	470.340	500.000	-5.9
Aroclor-1016-2	5.497	5.408	5.608	505.670	500.000	1.1
Aroclor-1016-3	5.559	5.470	5.670	559.330	500.000	11.9
Aroclor-1016-4	5.654	5.566	5.766	541.740	500.000	8.3
Aroclor-1016-5	5.946	5.858	6.058	522.030	500.000	4.4
Aroclor-1260-1	7.060	6.971	7.171	518.800	500.000	3.8
Aroclor-1260-2	7.313	7.225	7.425	539.800	500.000	8.0
Aroclor-1260-3	7.671	7.582	7.782	522.570	500.000	4.5
Aroclor-1260-4	7.893	7.803	8.003	501.910	500.000	0.4
Aroclor-1260-5	8.199	8.110	8.310	491.180	500.000	-1.8
Decachlorobiphenyl	9.911	9.830	10.030	58.250	50.000	16.5
Tetrachloro-m-xylene	4.336	4.248	4.448	56.580	50.000	13.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115552.D **Time Analyzed:** 09:34

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.714	4.628	4.828	483.410	500.000	-3.3
Aroclor-1016-2	4.734	4.647	4.847	463.460	500.000	-7.3
Aroclor-1016-3	4.908	4.821	5.021	430.720	500.000	-13.9
Aroclor-1016-4	4.950	4.864	5.064	406.600	500.000	-18.7
Aroclor-1016-5	5.162	5.076	5.276	538.120	500.000	7.6
Aroclor-1260-1	6.191	6.106	6.306	441.910	500.000	-11.6
Aroclor-1260-2	6.379	6.294	6.494	409.120	500.000	-18.2
Aroclor-1260-3	6.531	6.446	6.646	416.000	500.000	-16.8
Aroclor-1260-4	7.002	6.917	7.117	410.560	500.000	-17.9
Aroclor-1260-5	7.243	7.158	7.358	412.150	500.000	-17.6
Decachlorobiphenyl	8.623	8.541	8.741	53.610	50.000	7.2
Tetrachloro-m-xylene	3.640	3.548	3.748	52.170	50.000	4.3