

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115565.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.475	5.387	5.587	486.940	500.000	-2.6
Aroclor-1016-2	5.496	5.408	5.608	462.390	500.000	-7.5
Aroclor-1016-3	5.557	5.470	5.670	489.160	500.000	-2.2
Aroclor-1016-4	5.654	5.566	5.766	479.290	500.000	-4.1
Aroclor-1016-5	5.945	5.858	6.058	473.190	500.000	-5.4
Aroclor-1260-1	7.059	6.971	7.171	504.760	500.000	1.0
Aroclor-1260-2	7.313	7.225	7.425	514.490	500.000	2.9
Aroclor-1260-3	7.670	7.582	7.782	508.230	500.000	1.6
Aroclor-1260-4	7.893	7.803	8.003	477.910	500.000	-4.4
Aroclor-1260-5	8.199	8.110	8.310	468.510	500.000	-6.3
Decachlorobiphenyl	9.914	9.830	10.030	57.970	50.000	15.9
Tetrachloro-m-xylene	4.336	4.248	4.448	57.090	50.000	14.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/04/2025 11/04/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PO115565.D **Time Analyzed:** 14:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.715	4.628	4.828	488.550	500.000	-2.3
Aroclor-1016-2	4.734	4.647	4.847	467.850	500.000	-6.4
Aroclor-1016-3	4.908	4.821	5.021	469.280	500.000	-6.1
Aroclor-1016-4	4.949	4.864	5.064	415.460	500.000	-16.9
Aroclor-1016-5	5.162	5.076	5.276	547.590	500.000	9.5
Aroclor-1260-1	6.191	6.106	6.306	431.240	500.000	-13.8
Aroclor-1260-2	6.379	6.294	6.494	402.390	500.000	-19.5
Aroclor-1260-3	6.531	6.446	6.646	394.340	500.000	-21.1
Aroclor-1260-4	7.002	6.917	7.117	392.140	500.000	-21.6
Aroclor-1260-5	7.244	7.158	7.358	391.850	500.000	-21.6
Decachlorobiphenyl	8.624	8.541	8.741	50.710	50.000	1.4
Tetrachloro-m-xylene	3.640	3.548	3.748	53.110	50.000	6.2