



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3743</u>	Analytical Method:	<u>608.3 PCB</u>
Client:	<u>Remington & Vernick</u>	Datafile :	<u>PO115544.D</u>

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB170783BS (Column 1)	AROCLOR 1016	0.05	0.049	ug/L	97				50	140	
	AROCLOR 1260	0.05	0.052	ug/L	104				8	140	
PB170783BS (Column 2)	AROCLOR 1016	0.05	0.050	ug/L	99				50	140	
	AROCLOR 1260	0.05	0.048	ug/L	95				8	140	