

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3743

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/25/2025

11/25/2025

Continuing Calib Time: 12:54

Initial Calibration Time(s): 10:55

12:16

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	8.97	8.97	8.87	9.07	0.00
Tetrachloro-m-xylene	3.51	3.50	3.40	3.60	-0.01
alpha-BHC	3.95	3.95	3.85	4.05	0.00
beta-BHC	4.47	4.47	4.37	4.57	0.00
delta-BHC	4.72	4.71	4.61	4.81	-0.01
gamma-BHC (Lindane)	4.28	4.28	4.18	4.38	0.00
Heptachlor	4.87	4.87	4.77	4.97	0.00
Aldrin	5.21	5.21	5.11	5.31	0.00
Heptachlor epoxide	5.63	5.63	5.53	5.73	0.00
Endosulfan I	6.01	6.01	5.91	6.11	0.00
Dieldrin	6.28	6.28	6.18	6.38	0.00
4,4'-DDE	6.14	6.13	6.03	6.23	-0.01
Endrin	6.51	6.51	6.41	6.61	0.00
Endosulfan II	6.72	6.72	6.62	6.82	0.00
4,4'-DDD	6.65	6.64	6.54	6.74	0.00
Endosulfan sulfate	7.09	7.08	6.98	7.18	-0.01
4,4'-DDT	6.96	6.96	6.86	7.06	0.00
Methoxychlor	7.43	7.43	7.33	7.53	0.00
Endrin ketone	7.57	7.56	7.46	7.66	-0.01
Endrin aldehyde	6.85	6.85	6.75	6.95	0.00
alpha-Chlordane	5.96	5.96	5.86	6.06	0.00
gamma-Chlordane	5.88	5.88	5.78	5.98	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: REMI01

Lab Code: ACE

SDG NO.: Q3743

Continuing Calib Date: 12/02/2025

Initial Calibration Date(s): 11/25/2025

11/25/2025

Continuing Calib Time: 12:54

Initial Calibration Time(s): 10:55

12:16

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Decachlorobiphenyl	7.95	7.95	7.85	8.05	0.00
Tetrachloro-m-xylene	2.80	2.80	2.70	2.90	0.00
alpha-BHC	3.31	3.31	3.21	3.41	0.00
beta-BHC	3.93	3.93	3.83	4.03	0.00
delta-BHC	4.16	4.16	4.06	4.26	0.00
gamma-BHC (Lindane)	3.64	3.64	3.54	3.74	0.00
Heptachlor	3.98	3.98	3.88	4.08	0.00
Aldrin	4.26	4.26	4.16	4.36	0.00
Heptachlor epoxide	4.76	4.76	4.66	4.86	0.00
Endosulfan I	5.13	5.13	5.03	5.23	0.00
Dieldrin	5.40	5.40	5.30	5.50	0.00
4,4'-DDE	5.27	5.27	5.17	5.37	0.00
Endrin	5.67	5.67	5.57	5.77	0.00
Endosulfan II	5.97	5.96	5.86	6.06	-0.01
4,4'-DDD	5.82	5.82	5.72	5.92	0.00
Endosulfan sulfate	6.37	6.37	6.27	6.47	0.00
4,4'-DDT	6.07	6.07	5.97	6.17	0.00
Methoxychlor	6.65	6.64	6.54	6.74	-0.01
Endrin ketone	6.87	6.87	6.77	6.97	0.00
Endrin aldehyde	6.14	6.14	6.04	6.24	0.00
alpha-Chlordane	5.08	5.08	4.98	5.18	0.00
gamma-Chlordane	5.02	5.02	4.92	5.12	0.00