

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/25/2025 11/25/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097762.D **Time Analyzed:** 12:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	6.645	6.542	6.742	45.220	50.000	-9.6
4,4'-DDE	6.136	6.032	6.232	47.530	50.000	-4.9
4,4'-DDT	6.959	6.855	7.055	52.480	50.000	5.0
Aldrin	5.210	5.107	5.307	48.340	50.000	-3.3
alpha-BHC	3.952	3.850	4.050	48.230	50.000	-3.5
alpha-Chlordane	5.964	5.860	6.060	47.170	50.000	-5.7
beta-BHC	4.470	4.368	4.568	47.870	50.000	-4.3
Decachlorobiphenyl	8.973	8.866	9.066	43.120	50.000	-13.8
delta-BHC	4.715	4.613	4.813	50.360	50.000	0.7
Dieldrin	6.284	6.180	6.380	46.560	50.000	-6.9
Endosulfan I	6.011	5.908	6.108	48.070	50.000	-3.9
Endosulfan II	6.723	6.621	6.821	45.710	50.000	-8.6
Endosulfan sulfate	7.087	6.983	7.183	46.640	50.000	-6.7
Endrin	6.510	6.407	6.607	48.270	50.000	-3.5
Endrin aldehyde	6.853	6.750	6.950	46.510	50.000	-7.0
Endrin ketone	7.565	7.461	7.661	48.160	50.000	-3.7
gamma-BHC (Lindane)	4.281	4.179	4.379	46.450	50.000	-7.1
gamma-Chlordane	5.883	5.781	5.981	46.290	50.000	-7.4
Heptachlor	4.871	4.768	4.968	46.920	50.000	-6.2
Heptachlor epoxide	5.630	5.527	5.727	48.480	50.000	-3.0
Methoxychlor	7.432	7.328	7.528	50.020	50.000	0.0
Tetrachloro-m-xylene	3.505	3.404	3.604	45.470	50.000	-9.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** REMI01
Lab Code: ACE **SDG NO.:** Q3743
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/25/2025 11/25/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/02/2025
Lab Sample No.: PSTDCCC050 **Data File :** PL097762.D **Time Analyzed:** 12:54

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
4,4'-DDD	5.822	5.720	5.920	45.480	50.000	-9.0
4,4'-DDE	5.267	5.167	5.367	45.480	50.000	-9.0
4,4'-DDT	6.074	5.971	6.171	50.620	50.000	1.2
Aldrin	4.263	4.162	4.362	45.130	50.000	-9.7
alpha-BHC	3.305	3.205	3.405	44.960	50.000	-10.1
alpha-Chlordane	5.080	4.978	5.178	41.980	50.000	-16.0
beta-BHC	3.932	3.832	4.032	44.570	50.000	-10.9
Decachlorobiphenyl	7.954	7.850	8.050	45.510	50.000	-9.0
delta-BHC	4.164	4.063	4.263	45.160	50.000	-9.7
Dieldrin	5.399	5.297	5.497	46.580	50.000	-6.8
Endosulfan I	5.133	5.033	5.233	41.320	50.000	-17.4
Endosulfan II	5.966	5.863	6.063	45.890	50.000	-8.2
Endosulfan sulfate	6.367	6.265	6.465	46.200	50.000	-7.6
Endrin	5.673	5.571	5.771	45.220	50.000	-9.6
Endrin aldehyde	6.144	6.042	6.242	42.940	50.000	-14.1
Endrin ketone	6.871	6.768	6.968	46.120	50.000	-7.8
gamma-BHC (Lindane)	3.636	3.536	3.736	46.070	50.000	-7.9
gamma-Chlordane	5.016	4.915	5.115	44.770	50.000	-10.5
Heptachlor	3.982	3.881	4.081	45.220	50.000	-9.6
Heptachlor epoxide	4.763	4.663	4.863	47.850	50.000	-4.3
Methoxychlor	6.646	6.543	6.743	48.020	50.000	-4.0
Tetrachloro-m-xylene	2.801	2.702	2.902	44.280	50.000	-11.4