

Surrogate Summary

SDG No.: Q3743

Client: Remington & Vernick

Analytical Method: 608.3 Pest

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PL097735.D	PIBLK-PL097735.D	Tetrachloro-m-xyl	1	20	20.0	100		60	140
		Decachlorobiphen	1	20	21.4	107		60	140
		Tetrachloro-m-xyl	2	20	19.7	99		60	140
		Decachlorobiphen	2	20	21.0	105		60	140
I.BLK-PL097760.D	PIBLK-PL097760.D	Tetrachloro-m-xyl	1	20	23.9	119		60	140
		Decachlorobiphen	1	20	23.6	118		60	140
		Tetrachloro-m-xyl	2	20	22.3	112		60	140
		Decachlorobiphen	2	20	23.6	118		60	140
PB170784BL	PB170784BL	Tetrachloro-m-xyl	1	20	21.5	108		60	140
		Decachlorobiphen	1	20	21.0	105		60	140
		Tetrachloro-m-xyl	2	20	21.4	107		60	140
		Decachlorobiphen	2	20	23.0	115		60	140
PB170784BS	PB170784BS	Tetrachloro-m-xyl	1	20	22.1	110		60	140
		Decachlorobiphen	1	20	21.4	107		60	140
		Tetrachloro-m-xyl	2	20	20.3	101		60	140
		Decachlorobiphen	2	20	22.9	115		60	140
PB170784BSD	PB170784BSD	Tetrachloro-m-xyl	1	20	21.0	105		60	140
		Decachlorobiphen	1	20	21.9	110		60	140
		Tetrachloro-m-xyl	2	20	19.6	98		60	140
		Decachlorobiphen	2	20	22.9	115		60	140
Q3743-02	TW-1125-2	Tetrachloro-m-xyl	1	20	13.1	65		60	140
		Decachlorobiphen	1	20	16.7	83		60	140
		Tetrachloro-m-xyl	2	20	12.8	64		60	140
		Decachlorobiphen	2	20	35.9	179	*	60	140
I.BLK-PL097767.D	PIBLK-PL097767.D	Tetrachloro-m-xyl	1	20	23.9	119		60	140
		Decachlorobiphen	1	20	26.4	132		60	140
		Tetrachloro-m-xyl	2	20	23.0	115		60	140
		Decachlorobiphen	2	20	26.8	134		60	140