

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3743

Analytical Method: 608.3 Pest

Client: Remington & Vernick

Datafile : PL097764.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB170784BS (Column 1)	alpha-BHC	0.0125	0.0085	ug/L	68				37	140	
	gamma-BHC (Lindane)	0.0125	0.0092	ug/L	74				32	140	
	Heptachlor	0.0125	0.010	ug/L	83				34	140	
	Aldrin	0.0125	0.011	ug/L	86				42	140	
	beta-BHC	0.0125	0.010	ug/L	82				17	147	
	delta-BHC	0.0125	0.0025	ug/L	20				19	140	
	Heptachlor epoxide	0.0125	0.011	ug/L	89				37	142	
	Endosulfan I	0.0125	0.0091	ug/L	73				45	153	
	gamma-Chlordane	0.0125	0.012	ug/L	94				45	140	
	alpha-Chlordane	0.0125	0.010	ug/L	80				45	140	
	4,4'-DDE	0.0125	0.011	ug/L	91				30	145	
	Dieldrin	0.0125	0.011	ug/L	86				36	146	
	Endrin	0.0125	0.012	ug/L	97				30	147	
	Endosulfan II	0.0125	0.010	ug/L	82				1	202	
	4,4'-DDD	0.0125	0.0098	ug/L	78				31	141	
	4,4'-DDT	0.0125	0.011	ug/L	90				25	160	
	Endrin aldehyde	0.0125	0.011	ug/L	85				38	141	
	Endosulfan sulfate	0.0125	0.0088	ug/L	70				26	144	
	Methoxychlor	0.0125	0.011	ug/L	89				30	151	
	Endrin ketone	0.0125	0.011	ug/L	86				44	149	
PB170784BS (Column 2)	alpha-BHC	0.0125	0.010	ug/L	81				37	140	
	gamma-BHC (Lindane)	0.0125	0.010	ug/L	83				32	140	
	Heptachlor	0.0125	0.012	ug/L	95				34	140	
	Aldrin	0.0125	0.011	ug/L	85				42	140	
	beta-BHC	0.0125	0.012	ug/L	92				17	147	
	delta-BHC	0.0125	0.0027	ug/L	22				19	140	
	Heptachlor epoxide	0.0125	0.011	ug/L	90				37	142	
	Endosulfan I	0.0125	0.0093	ug/L	74				45	153	
	gamma-Chlordane	0.0125	0.011	ug/L	85				45	140	
	alpha-Chlordane	0.0125	0.011	ug/L	84				45	140	
	4,4'-DDE	0.0125	0.011	ug/L	84				30	145	
	Dieldrin	0.0125	0.012	ug/L	95				36	146	
	Endrin	0.0125	0.011	ug/L	88				30	147	
	Endosulfan II	0.0125	0.011	ug/L	90				1	202	
	4,4'-DDD	0.0125	0.0083	ug/L	66				31	141	
	4,4'-DDT	0.0125	0.011	ug/L	91				25	160	
	Endrin aldehyde	0.0125	0.0079	ug/L	63				38	141	
	Endosulfan sulfate	0.0125	0.0091	ug/L	73				26	144	
	Methoxychlor	0.0125	0.012	ug/L	96				30	151	
	Endrin ketone	0.0125	0.011	ug/L	88				44	149	