

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3743</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>12/03/2025</u> <u>10:58</u>
Lab File ID:	<u>VV039447.D</u>	Init. Calib. Date(s):	<u>12/03/2025</u> <u>12/03/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:55</u> <u>12:18</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRFAL3	MIN RRF	%D	MAX%D
Vinyl chloride	1.082	1.023	0.1	-5.4	30
Trichloroethene	0.606	0.585	0.3	-3.5	30
1,2-Dibromoethane	0.409	0.363	0.01	-11.4	20
1,2,3-Trichloropropane	0.511	0.487	0.01	-4.7	30
1,1,2,2-Tetrachloroethane	0.502	0.458	0.1	-8.6	20
1,2-Dibromo-3-chloropropane	0.107	0.108	0.01	1.5	40
Vinyl Chloride-d3	0.491	0.499	0.01	1.6	40
1,2-Dichloroethane-d4	0.333	0.341	0.01	2.4	20
1,2-Dichloropropane-d6	0.422	0.402	0.01	-4.6	20
Toluene-d8	0.712	0.616	0.01	-13.4	20
1,1,2,2-Tetrachloroethane-d2	0.238	0.249	0.01	4.7	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.