

Surrogate Summary

SDG No.: **Q3743**

Client: **Remington & Vernick**

Analytical Method: **SWSFAM_VOCSIM**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3743-01	TW-1125-1	Vinyl chloride-d3	0.5	0.57	114		40	130
		1,2-Dichloroethane-d4	0.5	0.45	90		70	130
		1,2-Dichloropropane-d6	0.5	0.42	85		60	140
		Toluene-d8	0.5	0.48	96		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.54	108		65	120
Q3743-02	TW-1125-2	Vinyl chloride-d3	0.5	0.50	100		40	130
		1,2-Dichloroethane-d4	0.5	0.43	85		70	130
		1,2-Dichloropropane-d6	0.5	0.41	82		60	140
		Toluene-d8	0.5	0.44	87		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.52	103		65	120
Q3743-05	TB	Vinyl chloride-d3	0.5	0.53	105		40	130
		1,2-Dichloroethane-d4	0.5	0.44	88		70	130
		1,2-Dichloropropane-d6	0.5	0.42	85		60	140
		Toluene-d8	0.5	0.48	96		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.53	106		65	120
VV1203WBL01	VBLK234	Vinyl chloride-d3	0.5	0.54	109		40	130
		1,2-Dichloroethane-d4	0.5	0.45	91		70	130
		1,2-Dichloropropane-d6	0.5	0.41	82		60	140
		Toluene-d8	0.5	0.47	93		70	130
		1,1,2,2-Tetrachloroethane-d2	0.5	0.53	106		65	120