284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance | Contract: REMI01

Lab Code: ACE | SDG No.: Q3743

Instrument ID: MSVOA_N | Calibration Date(s): 11/24/2025 11/24/2025

Heated Purge: (Y/N) N | Calibration Time(s): 12:38 14:35

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID: RRF001 = VN088344.D RRF005 = VN088345.D RRF020 = VN088346.D
RRF050 = VN088347.D RRF100 = VN088348.D RRF150 = VN088349.D

COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
Dichlorodifluoromethane	0.719	0.676	0.796	0.741	0.838	0.736	0.751	7.7
Chloromethane	0.911	0.800	0.805	0.801	0.759	0.764	0.807	6.8
Vinyl Chloride	0.804	0.844	0.856	0.812	0.837	0.772	0.821	3.8
Bromomethane		0.407	0.392	0.396	0.387	0.392	0.395	1.9
Chloroethane	0.555	0.503	0.496	0.505	0.502	0.485	0.508	4.8
Trichlorofluoromethane	1.234	1.170	1.168	1.092	1.201	1.055	1.153	5.8
1,1,2-Trichlorotrifluoroethane	0.660	0.635	0.639	0.564	0.633	0.553	0.614	7.2
1,1-Dichloroethene	0.722	0.620	0.601	0.568	0.595	0.554	0.610	9.8
Acetone	0.412	0.355	0.322	0.335	0.305	0.314	0.341	11.5
Carbon Disulfide	2.061	2.043	2.043	1.983	1.996	1.833	1.993	4.2
Methyl tert-butyl Ether	2.126	2.239	2.245	2.504	2.265	2.333	2.285	5.5
Methyl Acetate	1.072	1.048	0.957	1.049	0.972	0.985	1.014	4.7
Methylene Chloride	0.797	0.752	0.704	0.756	0.686	0.678	0.729	6.4
trans-1,2-Dichloroethene	0.752	0.668	0.676	0.686	0.647	0.626	0.676	6.4
1,1-Dichloroethane	1.389	1.431	1.348	1.427	1.321	1.286	1.367	4.3
Cyclohexane		1.450	1.264	1.133	1.237	1.109	1.239	10.9
2-Butanone	0.509	0.538	0.548	0.596	0.540	0.556	0.548	5.2
Carbon Tetrachloride	0.565	0.504	0.548	0.499	0.570	0.509	0.532	6
cis-1,2-Dichloroethene	0.797	0.801	0.794	0.829	0.747	0.744	0.785	4.2
Bromochloromethane	0.724	0.722	0.684	0.735	0.624	0.597	0.681	8.5
Chloroform	1.377	1.362	1.351	1.392	1.291	1.285	1.343	3.3
1,1,1-Trichloroethane	1.247	1.219	1.192	1.158	1.166	1.104	1.181	4.3
Methylcyclohexane	0.534	0.542	0.606	0.527	0.662	0.561	0.572	9.1
Benzene	1.658	1.555	1.609	1.559	1.559	1.477	1.570	3.9
1,2-Dichloroethane	0.583	0.614	0.596	0.615	0.597	0.587	0.599	2.3
Trichloroethene	0.382	0.385	0.383	0.357	0.371	0.348	0.371	4.1
1,2-Dichloropropane	0.426	0.402	0.402	0.406	0.396	0.381	0.402	3.6
Bromodichloromethane	0.602	0.586	0.586	0.589	0.583	0.566	0.586	2
4-Methyl-2-Pentanone	0.540	0.616	0.661	0.673	0.658	0.651	0.633	7.8
Toluene	0.829	0.885	0.940	0.913	0.960	0.911	0.906	5.1

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: Alliance Contract: REMI01

Lab Code: ACE SDG No.: Q3743

Instrument ID: MSVOA_N Calibration Date(s): 11/24/2025 11/24/2025

Heated Purge: (Y/N) N Calibration Time(s): 12:38 14:35

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID: RRF001 = VN088344.D RRF005 = VN088345.D RRF020 = VN088346.D
RRF050 = VN088347.D RRF100 = VN088348.D RRF150 = VN088349.D

COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	$\overline{RRF}$	% RSD
t-1,3-Dichloropropene	0.567	0.553	0.598	0.631	0.631	0.625	0.601	5.7
cis-1,3-Dichloropropene	0.590	0.618	0.628	0.662	0.654	0.640	0.632	4.2
1,1,2-Trichloroethane	0.350	0.359	0.364	0.353	0.341	0.336	0.351	3
2-Hexanone	0.359	0.301	0.406	0.436	0.442	0.450	0.399	14.7
Dibromochloromethane	0.384	0.386	0.406	0.409	0.410	0.404	0.400	2.9
1,2-Dibromoethane	0.304	0.367	0.364	0.369	0.367	0.359	0.355	7.2
Tetrachloroethene	0.426	0.388	0.413	0.377	0.407	0.367	0.396	5.7
Chlorobenzene	1.114	1.142	1.161	1.106	1.145	1.070	1.123	3
Ethyl Benzene	1.881	1.860	2.014	1.938	2.133	1.952	1.963	5.1
m/p-Xylenes	0.658	0.690	0.769	0.733	0.788	0.728	0.728	6.6
o-Xylene	0.670	0.619	0.722	0.696	0.749	0.708	0.694	6.5
Styrene	0.884	1.025	1.163	1.193	1.284	1.210	1.127	12.9
Bromoform	0.224	0.273	0.285	0.285	0.302	0.295	0.277	10.1
Isopropylbenzene	3.339	3.528	3.944	3.624	4.040	3.706	3.697	7.1
1,1,2,2-Tetrachloroethane	1.377	1.218	1.231	1.127	1.135	1.082	1.195	8.9
1,3-Dichlorobenzene	1.730	1.573	1.720	1.606	1.676	1.574	1.647	4.3
1,4-Dichlorobenzene	1.747	1.825	1.786	1.681	1.702	1.608	1.725	4.5
1,2-Dichlorobenzene	1.518	1.574	1.583	1.521	1.595	1.505	1.550	2.5
1,2-Dibromo-3-Chloropropane	0.284	0.284	0.284	0.288	0.300	0.300	0.290	2.8
1,2,4-Trichlorobenzene	0.998	0.864	0.901	0.889	0.955	0.917	0.921	5.3
1,2,3-Trichlorobenzene	0.942	0.850	0.888	0.875	0.930	0.891	0.896	3.8
1,2-Dichloroethane-d4		0.954	0.916	0.999	0.882	0.960	0.942	4.7
Dibromofluoromethane		0.361	0.350	0.350	0.334	0.337	0.346	3.2
Toluene-d8		1.249	1.292	1.285	1.308	1.313	1.289	2
4-Bromofluorobenzene		0.402	0.413	0.448	0.466	0.482	0.442	7.7

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.