

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3743</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>12/01/2025 09:58</u>
Lab File ID:	<u>VN088370.D</u>	Init. Calib. Date(s):	<u>11/24/2025 11/24/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:38 14:35</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.751	0.815		8.52	20
Chloromethane	0.807	0.858	0.1	6.32	20
Vinyl Chloride	0.821	0.851		3.65	20
Bromomethane	0.395	0.408		3.29	20
Chloroethane	0.508	0.525		3.35	20
Trichlorofluoromethane	1.153	1.195		3.64	20
1,1,2-Trichlorotrifluoroethane	0.614	0.636		3.58	20
1,1-Dichloroethene	0.610	0.607		-0.49	20
Acetone	0.341	0.269		-21.11	20
Carbon Disulfide	1.993	2.083		4.52	20
Methyl tert-butyl Ether	2.285	2.626		14.92	20
Methyl Acetate	1.014	1.108		9.27	20
Methylene Chloride	0.729	0.791		8.51	20
trans-1,2-Dichloroethene	0.676	0.724		7.1	20
1,1-Dichloroethane	1.367	1.456	0.1	6.51	20
Cyclohexane	1.239	1.285		3.71	20
2-Butanone	0.548	0.541		-1.28	20
Carbon Tetrachloride	0.532	0.584		9.77	20
cis-1,2-Dichloroethene	0.785	0.861		9.68	20
Bromochloromethane	0.681	0.672		-1.32	20
Chloroform	1.343	1.443		7.45	20
1,1,1-Trichloroethane	1.181	1.220		3.3	20
Methylcyclohexane	0.572	0.664		16.08	20
Benzene	1.570	1.738		10.7	20
1,2-Dichloroethane	0.599	0.676		12.85	20
Trichloroethene	0.371	0.415		11.86	20
1,2-Dichloropropane	0.402	0.453		12.69	20
Bromodichloromethane	0.586	0.670		14.33	20
4-Methyl-2-Pentanone	0.633	0.747		18.01	20
Toluene	0.906	1.034		14.13	20
t-1,3-Dichloropropene	0.601	0.701		16.64	20
cis-1,3-Dichloropropene	0.632	0.748		18.35	20
1,1,2-Trichloroethane	0.351	0.400		13.96	20
2-Hexanone	0.399	0.458		14.79	20
Dibromochloromethane	0.400	0.464		16	20
1,2-Dibromoethane	0.355	0.413		16.34	20
Tetrachloroethene	0.396	0.425		7.32	20
Chlorobenzene	1.123	1.242	0.3	10.6	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>REMI01</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3743</u>
Instrument ID:	<u>MSVOA_N</u>	Calibration Date/Time:	<u>12/01/2025</u> <u>09:58</u>
Lab File ID:	<u>VN088370.D</u>	Init. Calib. Date(s):	<u>11/24/2025</u> <u>11/24/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>12:38</u> <u>14:35</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.963	2.187		11.41	20
m/p-Xylenes	0.728	0.816		12.09	20
o-Xylene	0.694	0.778		12.1	20
Styrene	1.127	1.336		18.55	20
Bromoform	0.277	0.337	0.1	21.66	20
Isopropylbenzene	3.697	4.179		13.04	20
1,1,2,2-Tetrachloroethane	1.195	1.313	0.3	9.87	20
1,3-Dichlorobenzene	1.647	1.852		12.45	20
1,4-Dichlorobenzene	1.725	1.901		10.2	20
1,2-Dichlorobenzene	1.550	1.757		13.35	20
1,2-Dibromo-3-Chloropropane	0.290	0.317		9.31	20
1,2,4-Trichlorobenzene	0.921	1.042		13.14	20
1,2,3-Trichlorobenzene	0.896	0.989		10.38	20
1,2-Dichloroethane-d4	0.942	0.958		1.7	20
Dibromofluoromethane	0.346	0.377		8.96	20
Toluene-d8	1.289	1.356		5.2	20
4-Bromofluorobenzene	0.442	0.472		6.79	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.