



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN112425\
Data File : VN088343.D
Acq On : 24 Nov 2025 08:17
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N112425W.M
Title : SW846 8260
Last Update : Tue Nov 25 01:34:52 2025



AutoFind: Scans 1847, 1848, 1849; Background Corrected with Scan 1839

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.6	12249	PASS
75	95	30	60	56.6	29403	PASS
95	95	100	100	100.0	51987	PASS
96	95	5	9	6.9	3605	PASS
173	174	0.00	2	1.4	523	PASS
174	95	50	100	70.8	36808	PASS
175	174	5	9	8.4	3099	PASS
176	174	95	101	96.4	35493	PASS
177	176	5	9	6.4	2260	PASS

Average of 12.817 to 12.829 min.: VN088343.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
35.95	660	47.90	508	61.95	3309	72.95	3051
37.00	4014	49.00	3039	63.05	2784	74.00	10530
37.95	3391	50.00	12249	63.80	113	75.00	29403
38.95	1502	51.00	4245	64.00	255	75.95	2531
39.95	311	51.90	213	64.90	158	76.90	202
40.95	180	54.80	110	65.05	474	77.10	198
42.90	161	55.10	103	66.95	229	77.70	63
44.00	531	55.90	1233	68.00	6110	77.90	320
44.95	637	56.95	2258	68.95	5997	78.90	2897
46.85	771	59.95	755	69.95	751	79.90	742
47.10	213	61.00	3348	71.85	421	80.95	2778

Average of 12.817 to 12.829 min.: VN088343.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
81.90	605	96.00	3605	129.70	84	156.90	87
82.80	89	103.85	320	129.90	147	170.70	51
85.90	119	104.90	161	134.90	52	171.10	82
86.95	1933	105.70	75	137.00	60	171.90	617
87.95	2571	105.95	190	140.90	629	172.60	144
90.70	103	116.00	445	141.80	113	173.05	523
90.95	348	116.75	347	142.90	817	174.00	36808
92.05	1981	117.90	250	144.85	179	174.95	3099
92.95	2935	118.85	465	145.90	56	176.00	35493
94.00	6957	127.90	108	147.90	68	176.95	2260
95.00	51987	129.00	114	154.80	69	206.95	15

Average of 12.817 to 12.829 min.: VN088343.D\data.ms

BFB

Modified:subtracted

m/z	abund.	m/z	abund.	m/z	abund.	m/z	abund.
281.30	53						