

Surrogate Summary

SDG No.: **Q3743**

Client: **Remington & Vernick**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3743-01	TW-1125-1	1,2-Dichloroethane-d4	50	61.7	123		74	125
		Dibromofluoromethane	50	53.8	108		75	124
		Toluene-d8	50	50.2	100		86	113
		4-Bromofluorobenzene	50	53.4	107		77	121
Q3743-02	TW-1125-2	1,2-Dichloroethane-d4	50	61.2	122		74	125
		Dibromofluoromethane	50	53.0	106		75	124
		Toluene-d8	50	50.1	100		86	113
		4-Bromofluorobenzene	50	53.5	107		77	121
Q3743-03	TB	1,2-Dichloroethane-d4	50	60.1	120		74	125
		Dibromofluoromethane	50	52.4	105		75	124
		Toluene-d8	50	50.1	100		86	113
		4-Bromofluorobenzene	50	53.5	107		77	121
Q3743-04		1,2-Dichloroethane-d4	50	60.9	122		74	125
		Dibromofluoromethane	50	53.0	106		75	124
		Toluene-d8	50	49.8	100		86	113
		4-Bromofluorobenzene	50	53.2	106		77	121
VN1201WBL01	VN1201WBL01	1,2-Dichloroethane-d4	50	57.8	116		74	125
		Dibromofluoromethane	50	52.2	104		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	52.4	105		77	121
VN1201WBS01	VN1201WBS01	1,2-Dichloroethane-d4	50	56.0	112		74	125
		Dibromofluoromethane	50	55.1	110		75	124
		Toluene-d8	50	54.9	110		86	113
		4-Bromofluorobenzene	50	53.3	107		77	121
VN1201WBSD0	VN1201WBSD01	1,2-Dichloroethane-d4	50	52.2	104		74	125
		Dibromofluoromethane	50	52.3	105		75	124
		Toluene-d8	50	53.2	106		86	113
		4-Bromofluorobenzene	50	51.1	102		77	121