

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3743
 Lab File ID: VN088370.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	262500	8.20	457607	9.08	408098	11.84
UPPER LIMIT	525000	8.7	915214	9.582	816196	12.341
LOWER LIMIT	131250	7.7	228804	8.582	204049	11.341
EPA SAMPLE NO.						
TW-1125-1	150985	8.21	353552	9.08	347429	11.84
TW-1125-2	203666	8.21	472280	9.08	460828	11.84
TB	177560	8.21	408212	9.08	395290	11.84
TB	168109	8.21	386817	9.08	372510	11.84
VN1201WBL01	188076	8.20	435319	9.08	418422	11.84
VN1201WBS01	256678	8.21	487472	9.08	435019	11.84
VN1201WBSD01	257123	8.21	472668	9.08	413438	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.