

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: REMI01
 Lab Code: ACE SDG NO.: Q3743
 Lab File ID: VN088370.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	197460	13.765				
UPPER LIMIT	394920	14.265				
LOWER LIMIT	98730	13.265				
EPA SAMPLE NO.						
TW-1125-1	164080	13.77				
TW-1125-2	215787	13.77				
TB	183327	13.77				
TB	175948	13.77				
VN1201WBL01	198799	13.77				
VN1201WBS01	207303	13.77				
VN1201WBSD01	196888	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.