

Instrument ID: **MSVOA_L**

Daily Analysis Runlog For Sequence/QC Batch ID # VL101425

Review By	Semsettin Yesilyurt	Review On	10/14/2025 2:02:14 PM		
Supervise By	Maresh Dadoda	Supervise On	10/16/2025 4:41:37 AM		
SubDirectory	VL101425	HP Acquire Method	TO15AIR	HP Processing Method	VL101425AIR.M
STD. NAME		STD REF.#			
Tune/Reschk		AP2684			
Initial Calibration Stds		AP2687,AP2688,AP2689			
CCC		AP2687			
Internal Standard/PEM		AP2684			
ICV/I.BLK		AP2690			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	BFB	VL043103.D	14 Oct 2025 08:42	SY/MD	Ok
2	VSTDIC0.5	VL043104.D	14 Oct 2025 09:57	SY/MD	Not Ok
3	VSTDICCC010	VL043105.D	14 Oct 2025 10:31	SY/MD	Ok,M
4	VSTDIC002	VL043106.D	14 Oct 2025 11:27	SY/MD	Ok,M
5	VSTDIC001	VL043107.D	14 Oct 2025 12:01	SY/MD	Ok,M
6	VSTDIC0.5	VL043108.D	14 Oct 2025 12:34	SY/MD	Ok,M
7	VSTDIC0.1	VL043109.D	14 Oct 2025 13:08	SY/MD	Ok,M
8	VSTDIC0.03	VL043110.D	14 Oct 2025 13:42	SY/MD	Ok,M
9	VSTDIC015	VL043111.D	14 Oct 2025 14:18	SY/MD	Ok,M
10	VSTDICV010	VL043112.D	14 Oct 2025 14:51	SY/MD	Ok,M

M : Manual Integration