

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN112425

Review By	John Carlone	Review On	11/25/2025 9:36:43 AM
Supervise By	Maresh Dadoda	Supervise On	11/26/2025 1:52:19 PM
SubDirectory	VN112425	HP Acquire Method	HP Processing Method 82N112425W.M

STD. NAME	STD REF.#
Tune/Reschk	VP136378
Initial Calibration Stds	VP136444,VP136445,VP136446,VP136447,VP136448,VP136449
CCC	
Internal Standard/PEM	
ICV/I.BLK	VP136450
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088343.D	24 Nov 2025 08:17		JC\MD	Ok
2	VSTDICC001	VSTDICC001	VN088344.D	24 Nov 2025 12:38		JC\MD	Ok,M
3	VSTDICC005	VSTDICC005	VN088345.D	24 Nov 2025 13:11	COM.#74 Peak not detected in 1 and 5 PPB	JC\MD	Ok,M
4	VSTDICC020	VSTDICC020	VN088346.D	24 Nov 2025 13:32		JC\MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VN088347.D	24 Nov 2025 13:53	Method failed for Comp. #74	JC\MD	Ok,M
6	VSTDICC100	VSTDICC100	VN088348.D	24 Nov 2025 14:14		JC\MD	Ok,M
7	VSTDICC150	VSTDICC150	VN088349.D	24 Nov 2025 14:35		JC\MD	Ok,M
8	IBLK	IBLK	VN088350.D	24 Nov 2025 14:56		JC\MD	Ok
9	VSTDICV050	ICVVN112425	VN088351.D	24 Nov 2025 15:18		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory	VN120125
HP Acquire Method	HP Processing Method
	82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088369.D	01 Dec 2025 08:49		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088370.D	01 Dec 2025 09:58	pH#Lot#V12668	JC\MD	Ok,M
3	VN1201MBL01	VN1201MBL01	VN088371.D	01 Dec 2025 10:33		JC\MD	Ok
4	VN1201WBL01	VN1201WBL01	VN088372.D	01 Dec 2025 10:53		JC\MD	Ok
5	VN1201WBS01	VN1201WBS01	VN088373.D	01 Dec 2025 11:27		JC\MD	Ok,M
6	VN1201WBSD01	VN1201WBSD01	VN088374.D	01 Dec 2025 12:10		JC\MD	Ok,M
7	Q3715-02	MW6	VN088375.D	01 Dec 2025 12:30	vial A pH<2	JC\MD	Ok,M
8	Q3741-07	TB-1	VN088376.D	01 Dec 2025 12:51	vial A pH<2 TB	JC\MD	Ok
9	Q3716-01	MW5	VN088377.D	01 Dec 2025 13:11	vial A pH<2 Need 20X	JC\MD	Dilution
10	Q3716-03	MW4	VN088378.D	01 Dec 2025 13:32	Need Straight Run	JC\MD	Not Ok
11	Q3716-04	MW7	VN088379.D	01 Dec 2025 13:52	Need 200X	JC\MD	Dilution
12	Q3715-01	MW2	VN088380.D	01 Dec 2025 14:13	E flag of previous sample	JC\MD	ReRun
13	Q3716-02	MW3	VN088381.D	01 Dec 2025 14:33	vial A pH<2	JC\MD	Ok
14	Q3738-01	MW4	VN088382.D	01 Dec 2025 14:54	Run @ Lower Dilution	JC\MD	Not Ok
15	Q3738-03	Field	VN088383.D	01 Dec 2025 15:15	vial A pH<2 FB	JC\MD	Ok
16	Q3738-02	MW5	VN088384.D	01 Dec 2025 15:35	vial A pH<2	JC\MD	Ok,M
17	Q3745-01	STORAGE-BLANK-SO	VN088385.D	01 Dec 2025 15:56	vial A pH<2 Tics Contamination	JC\MD	ReRun
18	Q3745-02	STORAGE-BLANK-WA	VN088386.D	01 Dec 2025 16:16	vial A pH<2	JC\MD	Ok

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120125

Review By	Review On
Supervise By	Supervise On
SubDirectory VN120125	HP Acquire Method HP Processing Method 82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136451
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136452,VP136453

19	Q3745-03	STORAGE-BLANK-WA	VN088387.D	01 Dec 2025 16:37	vial A pH<2	JC\MD	Ok
20	Q3745-04	STORAGE-BLANK-SAM	VN088388.D	01 Dec 2025 16:58	vial A pH<2	JC\MD	Ok
21	Q3743-03	TB	VN088389.D	01 Dec 2025 17:18	vial A pH<2 TB	JC\MD	Ok
22	Q3743-04	TB	VN088390.D	01 Dec 2025 17:39	vial A pH<2 TB	JC\MD	Ok
23	Q3743-05	TB	VN088391.D	01 Dec 2025 18:00	SIM Method on login	JC\MD	Not Ok
24	Q3743-01	TW-1125-1	VN088392.D	01 Dec 2025 18:20	vial A pH<2	JC\MD	Ok
25	Q3743-02	TW-1125-2	VN088393.D	01 Dec 2025 18:41	vial A pH<2	JC\MD	Ok
26	VSTDCCC050	VSTDCCC050EC	VN088394.D	01 Dec 2025 19:01		JC\MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120225

Review By	Review On
Supervise By	Supervise On
SubDirectory	VN120225
HP Acquire Method	HP Processing Method
	82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136457
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136458,VP136459,VP136463,VP136464,VP136465,VP136466,VP136467,VP136468

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VN088395.D	02 Dec 2025 08:19		JC\MD	Ok
2	VSTDCCC050	VSTDCCC050	VN088396.D	02 Dec 2025 10:00	pH#Lot#V12668	JC\MD	Ok,M
3	VN1202WBL01	VN1202WBL01	VN088397.D	02 Dec 2025 10:20		JC\MD	Ok
4	VN1202MBL01	VN1202MBL01	VN088398.D	02 Dec 2025 10:40		JC\MD	Ok
5	VN1202WBS01	VN1202WBS01	VN088399.D	02 Dec 2025 11:01	Recovery failed	JC\MD	Not Ok
6	VN1202WBS02	VN1202WBS02	VN088400.D	02 Dec 2025 11:34		JC\MD	Ok,M
7	IBLK	IBLK	VN088401.D	02 Dec 2025 11:54		JC\MD	Ok
8	IBLK	IBLK	VN088402.D	02 Dec 2025 12:14		JC\MD	Ok
9	Q3738-01	MW4	VN088403.D	02 Dec 2025 12:35	vial A pH<2	JC\MD	Ok
10	Q3716-01DL	MW5DL	VN088404.D	02 Dec 2025 12:55	vial B pH<2	JC\MD	Ok,M
11	Q3716-04DL	MW7DL	VN088405.D	02 Dec 2025 13:16	vial B pH<2	JC\MD	Ok,M
12	Q3716-03	MW4	VN088406.D	02 Dec 2025 13:36	vial A pH<2	JC\MD	Ok,M
13	Q3745-01	STORAGE-BLANK-SO	VN088407.D	02 Dec 2025 13:56	vial B pH<2	JC\MD	Ok
14	Q3530-07 0.2PPB	LOD-MDL-WATER-01-(	VN088408.D	02 Dec 2025 15:02		JC\MD	Ok,M
15	Q3530-07 0.5PPB	LOD-MDL-WATER-01-(	VN088409.D	02 Dec 2025 15:22		JC\MD	Ok,M
16	Q3530-07 0.75PPB	LOD-MDL-WATER-01-(	VN088410.D	02 Dec 2025 15:42		JC\MD	Ok,M
17	Q3530-07 2.5PPB	LOD-MDL-WATER-01-(	VN088411.D	02 Dec 2025 16:03		JC\MD	Ok,M
18	Q3530-08 1.0PPB	LOQ-WATER-02-QT4-2	VN088412.D	02 Dec 2025 16:23		JC\MD	Ok,M

Instrument ID: MSVOA_N

Daily Analysis Runlog For Sequence/QC Batch ID # VN120225

Review By	Review On
Supervise By	Supervise On
SubDirectory	VN120225
HP Acquire Method	HP Processing Method
	82N112425W.M
STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP136457
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP136458,VP136459,VP136463,VP136464,VP136465,VP136466,VP136467,VP136468

19	Q3530-08 5.0PPB	LOQ-WATER-02-QT4-2	VN088413.D	02 Dec 2025 16:44		JC\MD	Ok,M
20	VSTDCCC050	VSTDCCC050EC	VN088414.D	02 Dec 2025 17:04		JC\MD	Ok,M

M : Manual Integration