

Surrogate Summary

SDG No.: **Q3745**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3745-01	STORAGE-BLANK-SOIL-REF-6	1,2-Dichloroethane-d4	50	61.2	122		74	125
		Dibromofluoromethane	50	49.2	98		75	124
		Toluene-d8	50	47.7	95		86	113
		4-Bromofluorobenzene	50	51.0	102		77	121
Q3745-02	STORAGE-BLANK-WATER-REF-5	1,2-Dichloroethane-d4	50	62.2	124		74	125
		Dibromofluoromethane	50	52.9	106		75	124
		Toluene-d8	50	50.5	101		86	113
		4-Bromofluorobenzene	50	56.6	113		77	121
Q3745-03	STORAGE-BLANK-WATER-REF-4	1,2-Dichloroethane-d4	50	59.2	118		74	125
		Dibromofluoromethane	50	51.7	103		75	124
		Toluene-d8	50	50.2	100		86	113
		4-Bromofluorobenzene	50	51.8	104		77	121
Q3745-04	STORAGE-BLANK-SAMPLE-MANAC	1,2-Dichloroethane-d4	50	59.0	118		74	125
		Dibromofluoromethane	50	53.4	107		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	53.0	106		77	121
VN1201WBL01	VN1201WBL01	1,2-Dichloroethane-d4	50	57.8	116		74	125
		Dibromofluoromethane	50	52.2	104		75	124
		Toluene-d8	50	50.0	100		86	113
		4-Bromofluorobenzene	50	52.4	105		77	121
VN1201WBS01	VN1201WBS01	1,2-Dichloroethane-d4	50	56.0	112		74	125
		Dibromofluoromethane	50	55.1	110		75	124
		Toluene-d8	50	54.9	110		86	113
		4-Bromofluorobenzene	50	53.3	107		77	121
VN1201WBSD0	VN1201WBSD01	1,2-Dichloroethane-d4	50	52.2	104		74	125
		Dibromofluoromethane	50	52.3	105		75	124
		Toluene-d8	50	53.2	106		86	113
		4-Bromofluorobenzene	50	51.1	102		77	121
VN1202WBL01	VN1202WBL01	1,2-Dichloroethane-d4	50	61.1	122		74	125
		Dibromofluoromethane	50	52.4	105		75	124
		Toluene-d8	50	50.1	100		86	113
		4-Bromofluorobenzene	50	51.8	104		77	121
VN1202WBS02	VN1202WBS02	1,2-Dichloroethane-d4	50	52.1	104		74	125
		Dibromofluoromethane	50	52.1	104		75	124
		Toluene-d8	50	50.5	101		86	113
		4-Bromofluorobenzene	50	49.3	99		77	121