

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3745
 Lab File ID: VN088370.D Date Analyzed: 12/01/2025
 Instrument ID: MSVOA_N Time Analyzed: 09:58
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	197460	13.765				
UPPER LIMIT	394920	14.265				
LOWER LIMIT	98730	13.265				
EPA SAMPLE NO.						
STORAGE-BLANK-WATER-REF-5	230697	13.77				
STORAGE-BLANK-WATER-REF-4	177360	13.77				
STORAGE-BLANK-SAMPLE-MANAGEMENT	171972	13.77				
VN1201WBL01	198799	13.77				
VN1201WBS01	207303	13.77				
VN1201WBSD01	196888	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.