

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CHEM02
 Lab Code: ACE SDG NO.: Q3745
 Lab File ID: VN088396.D Date Analyzed: 12/02/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:00
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	206439	13.764				
UPPER LIMIT	412878	14.264				
LOWER LIMIT	103220	13.264				
EPA SAMPLE NO.						
STORAGE-BLANK-SOIL-REF-6	222144	13.77				
VN1202WBL01	185686	13.77				
VN1202WBS02	188408	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.