

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: CHEM02

Lab Code: ACE

SDG NO.: Q3745

Lab File ID: VN088369.D

BFB Injection Date: 12/01/2025

Instrument ID: MSVOA_N

BFB Injection Time: 08:49

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	22.9
75	30.0 - 60.0% of mass 95	59.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.9 (1.2) 1
174	50.0 - 100.0% of mass 95	71
175	5.0 - 9.0% of mass 174	5.7 (8) 1
176	95.0 - 101.0% of mass 174	68.9 (97.1) 1
177	5.0 - 9.0% of mass 176	5.2 (7.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN088370.D	12/01/2025	09:58
VN1201WBL01	VN1201WBL01	VN088372.D	12/01/2025	10:53
VN1201WBS01	VN1201WBS01	VN088373.D	12/01/2025	11:27
VN1201WBSD01	VN1201WBSD01	VN088374.D	12/01/2025	12:10
STORAGE-BLANK-WATER-REF-5	Q3745-02	VN088386.D	12/01/2025	16:16
STORAGE-BLANK-WATER-REF-4	Q3745-03	VN088387.D	12/01/2025	16:37
STORAGE-BLANK-SAMPLE-MANAGEMENT	Q3745-04	VN088388.D	12/01/2025	16:58