

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CLOU03

Lab Code: ACE

SDG NO.: Q3751

Continuing Calib Date: 12/04/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 15:45

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR1 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |       | DIFF<br>RT |
|----------------------|------------|-----------|-----------|-------|------------|
|                      |            |           | FROM      | TO    |            |
| Aroclor-1248-1 (1)   | 5.77       | 5.78      | 5.68      | 5.88  | 0.01       |
| Aroclor-1248-2 (2)   | 6.04       | 6.05      | 5.95      | 6.15  | 0.01       |
| Aroclor-1248-3 (3)   | 6.24       | 6.25      | 6.15      | 6.35  | 0.01       |
| Aroclor-1248-4 (4)   | 6.64       | 6.65      | 6.55      | 6.75  | 0.01       |
| Aroclor-1248-5 (5)   | 6.68       | 6.69      | 6.59      | 6.79  | 0.01       |
| Tetrachloro-m-xylene | 4.61       | 4.63      | 4.53      | 4.73  | 0.02       |
| Decachlorobiphenyl   | 10.37      | 10.38     | 10.28     | 10.48 | 0.02       |

### CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: CLOU03

Lab Code: ACE

SDG NO.: Q3751

Continuing Calib Date: 12/04/2025

Initial Calibration Date(s): 11/21/2025

11/21/2025

Continuing Calib Time: 15:45

Initial Calibration Time(s): 10:44

19:39

GC Column: ZB-MR2 ID: 0.32 (mm)

| COMPOUND             | CCAL<br>RT | AVG<br>RT | RT WINDOW |      | DIFF<br>RT |
|----------------------|------------|-----------|-----------|------|------------|
|                      |            |           | FROM      | TO   |            |
| Aroclor-1248-1 (1)   | 4.85       | 4.86      | 4.76      | 4.96 | 0.01       |
| Aroclor-1248-2 (2)   | 5.08       | 5.09      | 4.99      | 5.19 | 0.01       |
| Aroclor-1248-3 (3)   | 5.12       | 5.13      | 5.03      | 5.23 | 0.01       |
| Aroclor-1248-4 (4)   | 5.29       | 5.30      | 5.20      | 5.40 | 0.01       |
| Aroclor-1248-5 (5)   | 5.69       | 5.69      | 5.59      | 5.79 | 0.01       |
| Tetrachloro-m-xylene | 3.77       | 3.78      | 3.68      | 3.88 | 0.01       |
| Decachlorobiphenyl   | 8.76       | 8.77      | 8.67      | 8.87 | 0.01       |