

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076758.D **Time Analyzed:** 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.771	5.667	5.867	498.120	500.000	-0.4
Aroclor-1016-2	5.792	5.687	5.887	504.450	500.000	0.9
Aroclor-1016-3	5.854	5.750	5.950	500.530	500.000	0.1
Aroclor-1016-4	5.951	5.847	6.047	501.070	500.000	0.2
Aroclor-1016-5	6.244	6.140	6.340	510.060	500.000	2.0
Aroclor-1260-1	7.361	7.256	7.456	500.150	500.000	0.0
Aroclor-1260-2	7.613	7.509	7.709	489.260	500.000	-2.1
Aroclor-1260-3	7.972	7.867	8.067	493.290	500.000	-1.3
Aroclor-1260-4	8.198	8.093	8.293	488.420	500.000	-2.3
Aroclor-1260-5	8.521	8.416	8.616	488.930	500.000	-2.2
Decachlorobiphenyl	10.369	10.265	10.465	49.550	50.000	-0.9
Tetrachloro-m-xylene	4.618	4.514	4.714	50.000	50.000	0.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL01 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076758.D **Time Analyzed:** 17:01

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.852	4.750	4.950	508.130	500.000	1.6
Aroclor-1016-2	4.869	4.767	4.967	516.310	500.000	3.3
Aroclor-1016-3	5.044	4.942	5.142	515.230	500.000	3.0
Aroclor-1016-4	5.087	4.984	5.184	484.990	500.000	-3.0
Aroclor-1016-5	5.299	5.197	5.397	552.140	500.000	10.4
Aroclor-1260-1	6.326	6.224	6.424	512.690	500.000	2.5
Aroclor-1260-2	6.513	6.412	6.612	519.620	500.000	3.9
Aroclor-1260-3	6.665	6.564	6.764	497.970	500.000	-0.4
Aroclor-1260-4	7.134	7.033	7.233	503.740	500.000	0.7
Aroclor-1260-5	7.375	7.273	7.473	500.140	500.000	0.0
Decachlorobiphenyl	8.767	8.666	8.866	49.010	50.000	-2.0
Tetrachloro-m-xylene	3.774	3.671	3.871	51.820	50.000	3.6