

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076773.D **Time Analyzed:** 22:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.766	5.667	5.867	465.250	500.000	-7.0
Aroclor-1016-2	5.788	5.687	5.887	470.570	500.000	-5.9
Aroclor-1016-3	5.850	5.750	5.950	468.280	500.000	-6.3
Aroclor-1016-4	5.947	5.847	6.047	470.850	500.000	-5.8
Aroclor-1016-5	6.240	6.140	6.340	471.200	500.000	-5.8
Aroclor-1260-1	7.357	7.256	7.456	472.690	500.000	-5.5
Aroclor-1260-2	7.610	7.509	7.709	467.230	500.000	-6.6
Aroclor-1260-3	7.967	7.867	8.067	460.200	500.000	-8.0
Aroclor-1260-4	8.193	8.093	8.293	462.150	500.000	-7.6
Aroclor-1260-5	8.517	8.416	8.616	459.120	500.000	-8.2
Decachlorobiphenyl	10.364	10.265	10.465	46.590	50.000	-6.8
Tetrachloro-m-xylene	4.614	4.514	4.714	46.540	50.000	-6.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL02 **Date Analyzed:** 12/03/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076773.D **Time Analyzed:** 22:14

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.848	4.750	4.950	466.600	500.000	-6.7
Aroclor-1016-2	4.865	4.767	4.967	484.680	500.000	-3.1
Aroclor-1016-3	5.040	4.942	5.142	480.910	500.000	-3.8
Aroclor-1016-4	5.083	4.984	5.184	444.350	500.000	-11.1
Aroclor-1016-5	5.294	5.197	5.397	532.260	500.000	6.5
Aroclor-1260-1	6.322	6.224	6.424	475.390	500.000	-4.9
Aroclor-1260-2	6.510	6.412	6.612	486.710	500.000	-2.7
Aroclor-1260-3	6.661	6.564	6.764	464.200	500.000	-7.2
Aroclor-1260-4	7.131	7.033	7.233	468.820	500.000	-6.2
Aroclor-1260-5	7.371	7.273	7.473	466.280	500.000	-6.7
Decachlorobiphenyl	8.763	8.666	8.866	46.330	50.000	-7.3
Tetrachloro-m-xylene	3.770	3.671	3.871	48.400	50.000	-3.2