

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076790.D **Time Analyzed:** 09:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.768	5.667	5.867	477.660	500.000	-4.5
Aroclor-1016-2	5.790	5.687	5.887	471.130	500.000	-5.8
Aroclor-1016-3	5.853	5.750	5.950	464.250	500.000	-7.2
Aroclor-1016-4	5.949	5.847	6.047	483.200	500.000	-3.4
Aroclor-1016-5	6.242	6.140	6.340	481.040	500.000	-3.8
Aroclor-1260-1	7.359	7.256	7.456	455.910	500.000	-8.8
Aroclor-1260-2	7.612	7.509	7.709	457.690	500.000	-8.5
Aroclor-1260-3	7.970	7.867	8.067	451.870	500.000	-9.6
Aroclor-1260-4	8.196	8.093	8.293	448.260	500.000	-10.3
Aroclor-1260-5	8.520	8.416	8.616	448.310	500.000	-10.3
Decachlorobiphenyl	10.370	10.265	10.465	45.850	50.000	-8.3
Tetrachloro-m-xylene	4.616	4.514	4.714	46.430	50.000	-7.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL03 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076790.D **Time Analyzed:** 09:10

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.848	4.750	4.950	464.000	500.000	-7.2
Aroclor-1016-2	4.866	4.767	4.967	476.010	500.000	-4.8
Aroclor-1016-3	5.041	4.942	5.142	476.400	500.000	-4.7
Aroclor-1016-4	5.083	4.984	5.184	458.680	500.000	-8.3
Aroclor-1016-5	5.295	5.197	5.397	464.080	500.000	-7.2
Aroclor-1260-1	6.323	6.224	6.424	461.020	500.000	-7.8
Aroclor-1260-2	6.510	6.412	6.612	463.470	500.000	-7.3
Aroclor-1260-3	6.662	6.564	6.764	454.190	500.000	-9.2
Aroclor-1260-4	7.131	7.033	7.233	454.670	500.000	-9.1
Aroclor-1260-5	7.372	7.273	7.473	451.220	500.000	-9.8
Decachlorobiphenyl	8.764	8.666	8.866	44.440	50.000	-11.1
Tetrachloro-m-xylene	3.770	3.671	3.871	47.830	50.000	-4.3