

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076804.D **Time Analyzed:** 15:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.768	5.667	5.867	436.080	500.000	-12.8
Aroclor-1016-2	5.790	5.687	5.887	426.820	500.000	-14.6
Aroclor-1016-3	5.852	5.750	5.950	430.500	500.000	-13.9
Aroclor-1016-4	5.949	5.847	6.047	461.600	500.000	-7.7
Aroclor-1016-5	6.242	6.140	6.340	428.440	500.000	-14.3
Aroclor-1260-1	7.359	7.256	7.456	416.430	500.000	-16.7
Aroclor-1260-2	7.612	7.509	7.709	419.120	500.000	-16.2
Aroclor-1260-3	7.970	7.867	8.067	416.450	500.000	-16.7
Aroclor-1260-4	8.196	8.093	8.293	413.430	500.000	-17.3
Aroclor-1260-5	8.519	8.416	8.616	410.690	500.000	-17.9
Decachlorobiphenyl	10.367	10.265	10.465	42.470	50.000	-15.1
Tetrachloro-m-xylene	4.615	4.514	4.714	42.000	50.000	-16.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3751
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 11/21/2025 11/21/2025

Client Sample No.: CCAL04 **Date Analyzed:** 12/04/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP076804.D **Time Analyzed:** 15:12

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.849	4.750	4.950	435.520	500.000	-12.9
Aroclor-1016-2	4.866	4.767	4.967	456.770	500.000	-8.6
Aroclor-1016-3	5.041	4.942	5.142	457.570	500.000	-8.5
Aroclor-1016-4	5.083	4.984	5.184	435.940	500.000	-12.8
Aroclor-1016-5	5.296	5.197	5.397	453.800	500.000	-9.2
Aroclor-1260-1	6.323	6.224	6.424	438.910	500.000	-12.2
Aroclor-1260-2	6.510	6.412	6.612	439.260	500.000	-12.1
Aroclor-1260-3	6.663	6.564	6.764	430.320	500.000	-13.9
Aroclor-1260-4	7.132	7.033	7.233	434.790	500.000	-13.0
Aroclor-1260-5	7.372	7.273	7.473	430.020	500.000	-14.0
Decachlorobiphenyl	8.764	8.666	8.866	42.350	50.000	-15.3
Tetrachloro-m-xylene	3.770	3.671	3.871	45.350	50.000	-9.3