

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3752

Project: Linden Generating Station - 402

Run No.: LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BS							
TPH	mg/L	20.0	18.0		90	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q3752

Project: Linden Generating Station - 402

Run No.: LB138151

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138151BSD							
TPH	mg/L	20.0	18.3		92	1	78-114	12/08/2025