

284 Sheffield Street, Mountainside, New Jersey 07092, Phone: 908 789 8900,

Fax: 908 789 8922

## **Report of Analysis**

Client: PSEG

Project: PSEG- Moorestown HQ

Client Sample ID: AR520-0002 Lab Sample ID: Q3753-05 Date Collected: 12/02/25 10:17

Date Received: 12/02/25 SDG No.: Q3753 Matrix: SOIL % Solid: 100

Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.
Corrosivity	6.28		1	0	0	рН		12/03/25 10:55	9045D
Ignitability	NO		1	0	0	oC		12/03/25 15:54	1030
Reactive Cyanide	0.017	J	1	0.0084	0.050	mg/Kg	12/02/25 15:30	12/03/25 11:32	9012B
Reactive Sulfide	4.73	J	1	0.20	10.0	mg/Kg	12/04/25 09:05	12/04/25 12:43	9034

Comments: pH result reported at temperature 20.7 °C

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

H = Sample Analysis Out Of Hold Time

J = Estimated Value

B = Analyte Found in Associated Method Blank

\* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N =Spiked sample recovery not within control limits