

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3756

Analytical Method: 8270-Modified

Client: JACOBS Engineering Group, Inc.

DataFile: BN038290.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB170838BSD	1,4-Dioxane	0.4	0.31	ug/L	78	6			20 (65)	160 (116)	20 (27)