

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3756

Client ID : SSTDCCC0.4

Date Analyzed: 12/05/2025

Lab File ID: BN038287.D

Time Analyzed: 23:42

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6906	7.761	20829	10.56	10600	14.42
UPPER LIMIT	13812	8.261	41658	11.062	21200	14.92
LOWER LIMIT	3453	7.261	10414.5	10.062	5300	13.92
EPA SAMPLE NO.						
01 PB170838BL	7136	7.76	19157	10.56	8410	14.43
02 PB170838BS	7065	7.76	18764	10.56	7890	14.42
03 PB170838BSD	6710	7.76	17808	10.56	7711	14.42
04 OW-04B-26.5-120225	6484	7.76	18611	10.56	9161	14.42
05 BR-01-190-120225	6614	7.76	18511	10.56	9412	14.42
06 MW-04-12-120225	6896	7.76	20829	10.56	11092	14.42
07 MW-13B-47.5-120225	7020	7.76	21534	10.56	11522	14.42
08 EB01-120225	6963	7.76	20031	10.56	11010	14.42

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.