

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3756

Client ID : SSTDCCC0.4

Date Analyzed: 12/11/2025

Lab File ID: BN038330.D

Time Analyzed: 23:59

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	5325	7.724	14404	10.52	7509	14.39
UPPER LIMIT	10650	8.224	28808	11.019	15018	14.887
LOWER LIMIT	2662.5	7.224	7202	10.019	3754.5	13.887
EPA SAMPLE NO.						
01 BR-01-190-120225DL	5762	7.72	15461	10.53	8101	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.