

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>JAC005</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3756</u>
Instrument ID:	<u>MSVOA_V</u>	Calibration Date/Time:	<u>12/05/2025</u> <u>16:50</u>
Lab File ID:	<u>VV039484.D</u>	Init. Calib. Date(s):	<u>12/03/2025</u> <u>12/03/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:55</u> <u>12:18</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRFCAL	MIN RRF	%D	MAX%D
Vinyl chloride	1.082	1.103	0.01	2	50
Vinyl Chloride-d3	0.491	0.570	0.01	16.2	50
1,2-Dichloroethane-d4	0.333	0.332	0.01	-0.4	50
1,2-Dichloropropane-d6	0.422	0.431	0.1	2.2	50
Toluene-d8	0.712	0.675	0.2	-5.1	50
1,1,2,2-Tetrachloroethane-d2	0.238	0.249	0.01	4.8	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.