

Surrogate Summary

SDG No.: **Q3756**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3756-01	OW-04B-26.5-120225	1,2-Dichloroethane-d4	50	56.4	113		70 (74)	130 (125)
		Dibromofluoromethane	50	46.7	93		70 (75)	130 (124)
		Toluene-d8	50	44.4	89		70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109		70 (77)	130 (121)
Q3756-03	BR-01-190-120225	1,2-Dichloroethane-d4	50	56.2	112		70 (74)	130 (125)
		Dibromofluoromethane	50	46.0	92		70 (75)	130 (124)
		Toluene-d8	50	44.0	88		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.7	105		70 (77)	130 (121)
Q3756-05	MW-04-12-120225	1,2-Dichloroethane-d4	50	55.2	110		70 (74)	130 (125)
		Dibromofluoromethane	50	47.1	94		70 (75)	130 (124)
		Toluene-d8	50	45.7	91		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.8	102		70 (77)	130 (121)
Q3756-07	MW-13B-47.5-120225	1,2-Dichloroethane-d4	50	55.7	111		70 (74)	130 (125)
		Dibromofluoromethane	50	47.2	94		70 (75)	130 (124)
		Toluene-d8	50	45.2	90		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.6	101		70 (77)	130 (121)
Q3756-09	EB01-120225	1,2-Dichloroethane-d4	50	58.9	118		70 (74)	130 (125)
		Dibromofluoromethane	50	50.4	101		70 (75)	130 (124)
		Toluene-d8	50	46.8	94		70 (86)	130 (113)
		4-Bromofluorobenzene	50	48.6	97		70 (77)	130 (121)
Q3756-10	TB01-120225	1,2-Dichloroethane-d4	50	65.5	131	*	70 (74)	130 (125)
		Dibromofluoromethane	50	51.8	104		70 (75)	130 (124)
		Toluene-d8	50	48.2	96		70 (86)	130 (113)
		4-Bromofluorobenzene	50	50.9	102		70 (77)	130 (121)
Q3756-10RE	TB01-120225RE	1,2-Dichloroethane-d4	50	54.6	109		70 (74)	130 (125)
		Dibromofluoromethane	50	44.8	90		70 (75)	130 (124)
		Toluene-d8	50	42.0	84		70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.1	88		70 (77)	130 (121)
VN1203WBL01	VN1203WBL01	1,2-Dichloroethane-d4	50	59.5	119		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	47.6	95		70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.1	94		70 (77)	130 (121)
VN1203WBS01	VN1203WBS01	1,2-Dichloroethane-d4	50	54.3	109		70 (74)	130 (125)
		Dibromofluoromethane	50	51.3	103		70 (75)	130 (124)
		Toluene-d8	50	49.5	99		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.4	99		70 (77)	130 (121)
VN1203WBSD0	VN1203WBSD01	1,2-Dichloroethane-d4	50	54.9	110		70 (74)	130 (125)
		Dibromofluoromethane	50	51.9	104		70 (75)	130 (124)
		Toluene-d8	50	52.0	104		70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.2	104		70 (77)	130 (121)
VN1204WBL01	VN1204WBL01	1,2-Dichloroethane-d4	50	61.7	123		70 (74)	130 (125)
		Dibromofluoromethane	50	50.8	102		70 (75)	130 (124)
		Toluene-d8	50	46.7	93		70 (86)	130 (113)
		4-Bromofluorobenzene	50	49.9	100		70 (77)	130 (121)
VN1204WBS01	VN1204WBS01	1,2-Dichloroethane-d4	50	53.8	108		70 (74)	130 (125)
		Dibromofluoromethane	50	50.7	101		70 (75)	130 (124)
		Toluene-d8	50	51.7	103		70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.4	109		70 (77)	130 (121)

