

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3756</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>JACOBS Engineering Group, Inc.</u>	Datafile :	<u>VN088419.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1203WBS01	Vinyl chloride	20	19.5	ug/L	98			70 (65)	130 (117)	
	1,1-Dichloroethene	20	18.9	ug/L	95			70 (74)	130 (110)	
	1,1-Dichloroethane	20	21.6	ug/L	108			70 (78)	130 (112)	
	cis-1,2-Dichloroethene	20	20.7	ug/L	104			70 (77)	130 (110)	
	1,1,1-Trichloroethane	20	20.7	ug/L	104			70 (80)	130 (108)	
	Benzene	20	20.3	ug/L	102			70 (82)	130 (109)	
	1,2-Dichloroethane	20	22.3	ug/L	112			70 (80)	130 (115)	
	Trichloroethene	20	19.2	ug/L	96			70 (77)	130 (113)	
	1,1,2-Trichloroethane	20	21.1	ug/L	106			70 (83)	130 (112)	
	Tetrachloroethene	20	16.5	ug/L	83			70 (67)	130 (123)	