

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3756</u>	Analytical Method:	<u>SW8260D</u>
Client:	<u>JACOBS Engineering Group, Inc.</u>	Datafile :	<u>VN088432.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN1203WBSD01	Vinyl chloride	20	20.5	ug/L	103	5		70 (65)	130 (117)	20 (19)
	1,1-Dichloroethene	20	21.2	ug/L	106	11		70 (74)	130 (110)	20 (20)
	1,1-Dichloroethane	20	21.6	ug/L	108	0		70 (78)	130 (112)	20 (20)
	cis-1,2-Dichloroethene	20	20.8	ug/L	104	0		70 (77)	130 (110)	20 (20)
	1,1,1-Trichloroethane	20	21.9	ug/L	110	6		70 (80)	130 (108)	20 (20)
	Benzene	20	20.2	ug/L	101	1		70 (82)	130 (109)	20 (15)
	1,2-Dichloroethane	20	22.4	ug/L	112	0		70 (80)	130 (115)	20 (20)
	Trichloroethene	20	19.0	ug/L	95	1		70 (77)	130 (113)	20 (15)
	1,1,2-Trichloroethane	20	21.4	ug/L	107	1		70 (83)	130 (112)	20 (20)
	Tetrachloroethene	20	16.7	ug/L	84	1		70 (67)	130 (123)	20 (15)