

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
Lab Code: ACE SDG NO.: Q3756
Lab File ID: VN088416.D Date Analyzed: 12/03/2025
Instrument ID: MSVOA_N Time Analyzed: 10:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	266209	8.21	462230	9.08	427792	11.84
UPPER LIMIT	532418	8.706	924460	9.583	855584	12.341
LOWER LIMIT	133105	7.706	231115	8.583	213896	11.341
EPA SAMPLE NO.						
EB01-120225	173573	8.21	400248	9.08	396303	11.84
TB01-120225	192200	8.21	468050	9.08	464764	11.84
TB01-120225RE	223956	8.21	518301	9.08	518638	11.84
VN1203WBL01	179578	8.21	416539	9.08	402638	11.84
VN1203WBS01	204614	8.21	382235	9.08	345174	11.84
VN1203WBSD01	250101	8.21	473571	9.08	423667	11.84

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.