

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3756
 Lab File ID: VN088437.D Date Analyzed: 12/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	259230	8.21	454552	9.08	431157	11.84
UPPER LIMIT	518460	8.706	909104	9.577	862314	12.341
LOWER LIMIT	129615	7.706	227276	8.577	215579	11.341
EPA SAMPLE NO.						
OW-04B-26.5-120225	243663	8.21	555543	9.08	546868	11.84
BR-01-190-120225	246275	8.21	564051	9.08	564375	11.84
MW-04-12-120225	206051	8.21	468908	9.08	455230	11.84
MW-13B-47.5-120225	206071	8.21	471902	9.08	469221	11.84
VN1204WBL01	236600	8.21	556948	9.08	555675	11.84
VN1204WBS01	303745	8.21	559564	9.08	491842	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.