

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3756
 Lab File ID: VN088416.D Date Analyzed: 12/03/2025
 Instrument ID: MSVOA_N Time Analyzed: 10:09
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	213345	13.765				
UPPER LIMIT	426690	14.265				
LOWER LIMIT	106673	13.265				
EPA SAMPLE NO.						
EB01-120225	182664	13.77				
TB01-120225	219360	13.77				
TB01-120225RE	245675	13.77				
VN1203WBL01	186090	13.77				
VN1203WBS01	173546	13.77				
VN1203WBSD01	214103	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.