

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3756
 Lab File ID: VN088437.D Date Analyzed: 12/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	210277	13.765				
UPPER LIMIT	420554	14.265				
LOWER LIMIT	105139	13.265				
EPA SAMPLE NO.						
OW-04B-26.5-120225	268660	13.76				
BR-01-190-120225	278056	13.77				
MW-04-12-120225	220606	13.77				
MW-13B-47.5-120225	222578	13.77				
VN1204WBL01	258091	13.77				
VN1204WBS01	237731	13.77				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.