

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3757

Client ID : SSTDCCC0.4

Date Analyzed: 12/11/2025

Lab File ID: BN038311.D

Time Analyzed: 11:13

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	6458	7.724	17315	10.53	8838	14.40
UPPER LIMIT	12916	8.224	34630	11.03	17676	14.898
LOWER LIMIT	3229	7.224	8657.5	10.03	4419	13.898
EPA SAMPLE NO.						
01 BR-03D-490-120325	7415	7.72	20111	10.53	10732	14.40
02 MW-14B-46-120325	5537	7.72	15261	10.53	7974	14.40
03 RMW-05B-90.5-120325	5204	7.72	14267	10.53	7400	14.40
04 OW-05B-72-120325	5450	7.72	14372	10.53	7355	14.40
05 EB01-120325	5459	7.72	14139	10.53	7229	14.40

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.