

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: JACO05

Lab Code: ACE

SDG NO.: Q3757

Lab File ID: VV039444.D

BFB Injection Date: 12/03/2025

Instrument ID: MSVOA_V

BFB Injection Time: 09:23

GC Column: DB-624UI ID: 0.18 (mm)

Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.4
75	30.0 - 60.0% of mass 95	48.8
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	1.7 (1.8) 1
174	50.0 - 100.0% of mass 95	94.2
175	5.0 - 9.0% of mass 174	6.9 (7.3) 1
176	95.0 - 101.0% of mass 174	91.7 (97.4) 1
177	5.0 - 9.0% of mass 176	6 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTD0.02269	VSTD0.0269	VV039445.D	12/03/2025	09:55
VSTD0.05270	VSTD0.0570	VV039446.D	12/03/2025	10:32
VSTD0.1271	VSTD0.171	VV039447.D	12/03/2025	10:58
VSTD0.5272	VSTD0.572	VV039448.D	12/03/2025	11:25
VSTD001273	VSTD00173	VV039449.D	12/03/2025	11:52
VSTD002274	VSTD00274	VV039450.D	12/03/2025	12:18