

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JACO05
 Lab Code: ACE SDG NO.: Q3757
 Lab File ID: VN088437.D Date Analyzed: 12/04/2025
 Instrument ID: MSVOA_N Time Analyzed: 11:44
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	259230	8.21	454552	9.08	431157	11.84
UPPER LIMIT	518460	8.706	909104	9.577	862314	12.341
LOWER LIMIT	129615	7.706	227276	8.577	215579	11.341
EPA SAMPLE NO.						
EB01-120325	201753	8.21	451562	9.08	439722	11.84
TB01-120325	198228	8.21	454339	9.08	441078	11.84
VN1204WBL01	236600	8.21	556948	9.08	555675	11.84
VN1204WBS01	303745	8.21	559564	9.08	491842	11.84
VN1204WBSD01	290727	8.21	529277	9.08	469866	11.84

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.