

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3757
 Lab File ID: VX048723.D Date Analyzed: 12/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	218600	5.53	333007	6.74	285113	10.04
UPPER LIMIT	437200	6.025	666014	7.239	570226	10.537
LOWER LIMIT	109300	5.025	166504	6.239	142557	9.537
EPA SAMPLE NO.						
MW-14A-17.5-120325	176619	5.53	302604	6.74	274042	10.04
BR-03D-490-120325	163180	5.54	280929	6.75	230844	10.04
MW-14B-46-120325	161591	5.54	286549	6.75	220300	10.04
RMW-05B-90.5-120325	132392	5.54	222932	6.75	207009	10.04
OW-05B-72-120325	129477	5.54	275147	6.75	208806	10.04
VX1205WBL01	193187	5.53	291151	6.74	294164	10.04
VX1205WBS01	179194	5.53	293931	6.74	269397	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.