

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3757
 Lab File ID: VX048742.D Date Analyzed: 12/08/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:20
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	174813	5.53	271670	6.74	259893	10.04
UPPER LIMIT	349626	6.031	543340	7.239	519786	10.537
LOWER LIMIT	87406.5	5.031	135835	6.239	129947	9.537
EPA SAMPLE NO.						
MW-14B-46-120325RE	124903	5.54	299763	6.75	268328	10.04
OW-05B-72-120325RE	169111	5.54	409831	6.74	413547	10.04
VX1208WBL01	173343	5.53	340953	6.74	370451	10.04
VX1208WBS01	178446	5.53	298973	6.74	311271	10.04

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.