

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: JAC005
 Lab Code: ACE SDG NO.: Q3757
 Lab File ID: VX048723.D Date Analyzed: 12/05/2025
 Instrument ID: MSVOA_X Time Analyzed: 08:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	141895	12				
UPPER LIMIT	283790	12.5				
LOWER LIMIT	70947.5	11.5				
EPA SAMPLE NO.						
MW-14A-17.5-120325	141373	12.01				
BR-03D-490-120325	110489	12.01				
MW-14B-46-120325	131055	12.01				
RMW-05B-90.5-120325	102583	12.01				
OW-05B-72-120325	103213	12.01				
VX1205WBL01	130501	12.01				
VX1205WBS01	132992	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.