

## Initial and Continuing Calibration Verification

**Client:** Tully Environmental, Inc

**SDG No.:** Q3760

**Project:** Transfer Station-SPDES

**RunNo.:** LB138125

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: <b>ICV1</b> Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: <b>CCV1</b> Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: <b>CCV2</b> Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: <b>CCV3</b> Ammonia as N	mg/L	1	1	100	90-110	12/05/2025
Sample ID: <b>CCV4</b> Ammonia as N	mg/L	1	1	100	90-110	12/05/2025