

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3760

Project: Transfer Station-SPDES

Run No.: LB138103

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138103BS							
BOD5	mg/L	198	217		109	1	84.6-115.4	12/04/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3760

Project: Transfer Station-SPDES

Run No.: LB138118

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138118BS							
TSS	mg/L	550	556		101	1	90-110	12/04/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3760

Project: Transfer Station-SPDES

Run No.: LB138150

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB138150BS							
Oil and Grease	mg/L	20.0	16.7		84	1	78-114	12/08/2025

Laboratory Control Sample Summary

Client: Tully Environmental, Inc

SDG No.: Q3760

Project: Transfer Station-SPDES

Run No.: LB138125

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	PB170830BS							
Ammonia as N	mg/L	1	1.00		100	1	90-110	12/05/2025