

DATA REPORTING QUALIFIERS- INORGANIC

For reporting results, the following “ Results Qualifiers” are used:

J	Indicates the reported value was obtained from a reading that was less than the Contract Required Detection Limit (CRDL), but greater than or equal to the Instrument Detection Limit (IDL).
U	Indicates the analyte was analyzed for, but not detected.
ND	Indicates the analyte was analyzed for, but not detected
E	Indicates the reported value is estimated because of the presence of interference
M	Indicates Duplicate injection precision not met.
N	Indicates the spiked sample recovery is not within control limits.
S	Indicates the reported value was determined by the Method of Standard Addition (MSA).
*	Indicates that the duplicate analysis is not within control limits.
+	Indicates the correlation coefficient for the MSA is less than 0.995.
D	Indicates the reported value is from a secondary analysis with a dilution factor. The original analysis exceeded the calibration range.
M	Method qualifiers “P” for ICP instrument “PM” for ICP when Microwave Digestion is used “CV” for Manual Cold Vapor AA “AV” for automated Cold Vapor AA “CA” for MIDI-Distillation Spectrophotometric “AS” for Semi -Automated Spectrophotometric “C” for Manual Spectrophotometric “T” for Titrimetric “NR” for analyte not required to be analyzed
OR	Indicates the analyte’s concentration exceeds the calibrated range of the instrument for that specific analysis.
Q	Indicates the LCS did not meet the control limits requirements
H	Sample Analysis Out Of Hold Time

LAB CHRONICLE

OrderID:	Q3760	OrderDate:	12/3/2025 10:53:00 AM
Client:	Tully Environmental, Inc	Project:	Transfer Station-SPDES
Contact:	Dean Devoe	Location:	A11,VOA Lab

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3760-01	001 Willets Pt Blvd (Dec)	Water			12/02/25			12/03/25
			Metals Group 10	200.7		12/04/25	12/10/25	
Q3760-02	002 35th Ave (Dec)	Water			12/02/25			12/03/25
			Metals Group 10	200.7		12/04/25	12/10/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3760

Order ID: Q3760

Client: Tully Environmental, Inc

Project ID: Transfer Station-SPDES

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID : 001 Willets Pt Blvd (Dec)								
Q3760-01	001 Willets Pt Blvd (Dec)	Water	Copper	32.4		1.89	10.0	ug/L
Q3760-01	001 Willets Pt Blvd (Dec)	Water	Iron	7330		18.0	50.0	ug/L
Q3760-01	001 Willets Pt Blvd (Dec)	Water	Lead	31.5		1.21	6.00	ug/L
Client ID : 002 35th Ave (Dec)								
Q3760-02	002 35th Ave (Dec)	Water	Copper	28.4		1.89	10.0	ug/L
Q3760-02	002 35th Ave (Dec)	Water	Iron	6510		18.0	50.0	ug/L
Q3760-02	002 35th Ave (Dec)	Water	Lead	27.7		1.21	6.00	ug/L



SAMPLE DATA

Report of Analysis

Client: Tully Environmental, Inc
Project: Transfer Station-SPDES
Client Sample ID: 001 Willets Pt Blvd (Dec)
Lab Sample ID: Q3760-01
Level (low/med): low

Date Collected: 12/02/25
Date Received: 12/03/25
SDG No.: Q3760
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	32.4		1	1.89	10.0	ug/L	12/04/25 10:20	12/10/25 15:20	EPA 200.7	M200.7
7439-89-6	Iron	7330		1	18.0	50.0	ug/L	12/04/25 10:20	12/10/25 15:20	EPA 200.7	M200.7
7439-92-1	Lead	31.5		1	1.21	6.00	ug/L	12/04/25 10:20	12/10/25 15:20	EPA 200.7	M200.7

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits

Report of Analysis

Client: Tully Environmental, Inc
Project: Transfer Station-SPDES
Client Sample ID: 002 35th Ave (Dec)
Lab Sample ID: Q3760-02
Level (low/med): low

Date Collected: 12/02/25
Date Received: 12/03/25
SDG No.: Q3760
Matrix: Water
% Solid: 0

Cas	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Prep Date	Date Ana.	Ana Met.	Prep Met.
7440-50-8	Copper	28.4		1	1.89	10.0	ug/L	12/04/25 10:20	12/10/25 15:24	EPA 200.7	M200.7
7439-89-6	Iron	6510		1	18.0	50.0	ug/L	12/04/25 10:20	12/10/25 15:24	EPA 200.7	M200.7
7439-92-1	Lead	27.7		1	1.21	6.00	ug/L	12/04/25 10:20	12/10/25 15:24	EPA 200.7	M200.7

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

D = Dilution

Q = indicates LCS control criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

* = indicates the duplicate analysis is not within control limits.

E = Indicates the reported value is estimated because of the presence of interference.

OR = Over Range

N = Spiked sample recovery not within control limits



METAL CALIBRATION DATA

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	986	1000	99	95 - 105	P	12/10/2025	13:06	LB138187
	Iron	3840	4000	96	95 - 105	P	12/10/2025	13:06	LB138187
	Lead	4030	4000	101	95 - 105	P	12/10/2025	13:06	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.5	20.0	107	80 - 120	P	12/10/2025	13:16	LB138187
	Iron	105	100	105	80 - 120	P	12/10/2025	13:16	LB138187
	Lead	12.2	12.0	102	80 - 120	P	12/10/2025	13:16	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	12/10/2025	13:55	LB138187
	Iron	5080	5000	102	90 - 110	P	12/10/2025	13:55	LB138187
	Lead	4970	5000	99	90 - 110	P	12/10/2025	13:55	LB138187
CCV02	Copper	1250	1250	100	90 - 110	P	12/10/2025	14:45	LB138187
	Iron	5130	5000	103	90 - 110	P	12/10/2025	14:45	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	14:45	LB138187
CCV03	Copper	1260	1250	101	90 - 110	P	12/10/2025	15:46	LB138187
	Iron	5040	5000	101	90 - 110	P	12/10/2025	15:46	LB138187
	Lead	4930	5000	99	90 - 110	P	12/10/2025	15:46	LB138187
CCV04	Copper	1240	1250	99	90 - 110	P	12/10/2025	16:42	LB138187
	Iron	4920	5000	98	90 - 110	P	12/10/2025	16:42	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	16:42	LB138187
CCV05	Copper	1240	1250	99	90 - 110	P	12/10/2025	17:49	LB138187
	Iron	5010	5000	100	90 - 110	P	12/10/2025	17:49	LB138187
	Lead	4860	5000	97	90 - 110	P	12/10/2025	17:49	LB138187
CCV06	Copper	1220	1250	98	90 - 110	P	12/10/2025	18:41	LB138187
	Iron	4780	5000	96	90 - 110	P	12/10/2025	18:41	LB138187
	Lead	4820	5000	96	90 - 110	P	12/10/2025	18:41	LB138187
CCV07	Copper	1220	1250	98	90 - 110	P	12/10/2025	19:33	LB138187
	Iron	4800	5000	96	90 - 110	P	12/10/2025	19:33	LB138187
	Lead	4790	5000	96	90 - 110	P	12/10/2025	19:33	LB138187
CCV08	Copper	1240	1250	99	90 - 110	P	12/10/2025	20:21	LB138187
	Iron	4860	5000	97	90 - 110	P	12/10/2025	20:21	LB138187
	Lead	4870	5000	98	90 - 110	P	12/10/2025	20:21	LB138187



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Copper	21.4	20.0	107	65 - 135	P	12/10/2025	13:33	LB138187
	Iron	105	100	104	65 - 135	P	12/10/2025	13:33	LB138187
	Lead	13.7	12.0	114	65 - 135	P	12/10/2025	13:33	LB138187



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Copper	4.60	+/-10	U	20.0	P	12/10/2025	13:21	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	13:21	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:21	LB138187

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Copper	4.60	+/-10	U	20.0	P	12/10/2025	13:59	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	13:59	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	13:59	LB138187
CCB02	Copper	4.60	+/-10	U	20.0	P	12/10/2025	14:49	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	14:49	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	14:49	LB138187
CCB03	Copper	4.60	+/-10	U	20.0	P	12/10/2025	15:50	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	15:50	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	15:50	LB138187
CCB04	Copper	4.60	+/-10	U	20.0	P	12/10/2025	16:46	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	16:46	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	16:46	LB138187
CCB05	Copper	4.60	+/-10	U	20.0	P	12/10/2025	17:53	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	17:53	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	17:53	LB138187
CCB06	Copper	4.60	+/-10	U	20.0	P	12/10/2025	18:45	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	18:45	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	18:45	LB138187
CCB07	Copper	4.60	+/-10	U	20.0	P	12/10/2025	19:37	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	19:37	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	19:37	LB138187
CCB08	Copper	4.60	+/-10	U	20.0	P	12/10/2025	20:25	LB138187
	Iron	23.4	+/-50	U	100	P	12/10/2025	20:25	LB138187
	Lead	2.30	+/-6	U	12.0	P	12/10/2025	20:25	LB138187

Metals
- 3b -
PREPARATION BLANK SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Instrument: P4

Sample ID	Analyte	Result (ug/L)	Acceptance Limit	Conc Qual	CRQL ug/L	M	Analysis Date	Analysis Time	Run
PB170816BL	WATER			Batch Number:	PB170816		Prep Date:	12/04/2025	
	Copper	1.89	<5	U	10.0	P	12/10/2025	15:29	LB138187
	Iron	18.0	<25	U	50.0	P	12/10/2025	15:29	LB138187
	Lead	1.21	<3	U	6.00	P	12/10/2025	15:29	LB138187

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client:	<u>Tully Environmental, Inc</u>	SDG No.:	<u>Q3760</u>
Contract:	<u>TULL01</u>	Lab Code:	<u>ACE</u>
ICS Source:	<u>EPA</u>	Instrument ID:	<u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Copper	21.3	2.0	1065	-18	22	12/10/2025	13:38	LB138187
	Iron	96400	101000	95	85600	116500	12/10/2025	13:38	LB138187
	Lead	8.91			-12	12	12/10/2025	13:38	LB138187
ICSAB01	Copper	475	511	93	434	588	12/10/2025	13:42	LB138187
	Iron	91100	99300	92	84400	114500	12/10/2025	13:42	LB138187
	Lead	59.8	49.0	122	37	61	12/10/2025	13:42	LB138187
ICSA	Copper	23.6	2.0	1180	-18	22	12/10/2025	13:46	LB138187
	Iron	95500	101000	95	85600	116500	12/10/2025	13:46	LB138187
	Lead	46.9			-12	12	12/10/2025	13:46	LB138187
ICSAB	Copper	514	511	101	434	588	12/10/2025	13:50	LB138187
	Iron	90600	99300	91	84400	114500	12/10/2025	13:50	LB138187
	Lead	72.6	49.0	148	37	61	12/10/2025	13:50	LB138187



METAL

QC

DATA

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3760</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3771-02</u>	client id: <u>CompMS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3771-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	218		85.1		150	89		P
Iron	ug/L	75 - 125	8010		7140		1500	58		P
Lead	ug/L	75 - 125	451		9.00		500	88		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3760</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3771-02</u>	client id: <u>CompMSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3771-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	217		85.1		150	88		P
Iron	ug/L	75 - 125	8190		7140		1500	70		P
Lead	ug/L	75 - 125	454		9.00		500	89		P

Metals

- 5b -

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Matrix: _____

Level: LOW

Client ID: _____

Sample ID: _____

Spiked ID: _____

Analyte	Units	Acceptance Limit %R	C	Sample Result	C	Spike Added	% Recovery	Qual	M
---------	-------	------------------------	---	------------------	---	----------------	---------------	------	---

.....

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tully Environmental, Inc</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3760</u>
Contract:	<u>TULL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3771-02</u>	Client ID:	<u>CompDUP</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3771-02DUP	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	85.1		85.2		0		P
Iron	ug/L	20	7140		7050		1		P
Lead	ug/L	20	9.00		9.56		6		P

Metals

- 6 -

DUPLICATE SAMPLE SUMMARY

Client:	<u>Tully Environmental, Inc</u>	Level:	<u>LOW</u>	SDG No.:	<u>Q3760</u>
Contract:	<u>TULL01</u>			Lab Code:	<u>ACE</u>
Matrix:	<u>Water</u>	Sample ID:	<u>Q3771-02MS</u>	Client ID:	<u>CompMSD</u>
Percent Solids for Sample:	NA	Duplicate ID	Q3771-02MSD	Percent Solids for Spike Sample:	NA

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Copper	ug/L	20	218		217		0		P
Iron	ug/L	20	8010		8190		2		P
Lead	ug/L	20	451		454		1		P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170816BS							
Copper	ug/L	150	154		103	85 - 115	P
Iron	ug/L	1500	1470		98	85 - 115	P
Lead	ug/L	500	475		95	85 - 115	P

Metals

-9 -

ICP SERIAL DILUTIONS

SAMPLE NO.

CompL

Lab Name: Alliance

Contract: TULL01

Lab Code: ACE

Lb No.: lb138187

Lab Sample ID : Q3771-02L

SDG No.: Q3760

Matrix (soil/water): Water

Level (low/med): LOW

Concentration Units: ug/L

Analyte	Initial Sample Result (I) C		Serial Dilution Result (S) C		% Differ- ence	Q	M
Copper	85.1		89.2		5		P
Iron	7140		6530		8		P
Lead	9.00		7.78	J	14		P



METAL PREPARATION & INSTRUMENT DATA

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Al	Ca	Fe	Mg	Ag
Copper	224.700	0.0000000	0.0000000	0.0007850	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	-0.0000920	0.0000000	0.0000380	0.0000000	0.0000000

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Instrument ID: _____

Date: _____

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		As	Ba	Be	Cd	Co
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0000000	0.0009530
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	-0.0039600
Lead	220.353	0.0000000	0.0003170	0.0000000	0.0000000	0.0000000



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Cr	Cu	K	Mn	Mo
Copper	224.700	0.0000000	0.0000000	0.0000000	0.0006510	0.0020500
Iron	240.488	0.0000000	0.0000000	0.0000730	0.0000000	-0.0015250
Lead	220.353	0.0000000	0.0000000	0.0000000	0.0001400	-0.0008600



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interelement Correction Factors For:				
		Na	Ni	Pb	Sb	Se
Copper	224.700	0.0000000	-0.0047000	0.0036100	0.0000000	0.0000000
Iron	240.488	0.0000000	-0.0017000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	0.0006580	0.0000000	0.0000000	0.0001290



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Metals

- 11 -

ICP INTERELEMENT CORRECTION FACTORS

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Instrument ID:

Date:

Interelement Correction Factors (apparent ppb analyte/ppm interferent)

Analyte	Wave- Length (nm)	ICP Interement Correction Factors For:				
		Sn	Ti	Tl	V	Zn
Copper	224.700	0.0000000	0.0003840	0.0000000	0.0000000	0.0000000
Iron	240.488	0.0000000	0.0000000	0.0000000	0.0000000	0.0000000
Lead	220.353	0.0000000	-0.0003610	0.0000000	0.0000000	0.0000000



METAL PREPARATION & ANALYICAL SUMMARY

Metals

- 13 -

SAMPLE PREPARATION SUMMARY

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Method:

Sample ID	Client ID	Sample Type	Matrix	Prep Date	Initial Sample Size(mL)	Final Sample Volume (mL)	Percent Solids
Batch Number: PB170816							
PB170816BL	PB170816BL	MB	WATER	12/04/2025	50.0	25.0	
PB170816BS	PB170816BS	LCS	WATER	12/04/2025	50.0	25.0	
Q3760-01	001 Willets Pt Blvd (Dec)	SAM	WATER	12/04/2025	50.0	25.0	
Q3760-02	002 35th Ave (Dec)	SAM	WATER	12/04/2025	50.0	25.0	
Q3771-02DUP	CompDUP	DUP	WATER	12/04/2025	50.0	25.0	
Q3771-02MS	CompMS	MS	WATER	12/04/2025	50.0	25.0	
Q3771-02MSD	CompMSD	MSD	WATER	12/04/2025	50.0	25.0	

metals
- 14 -
ANALYSIS RUN LOG

Client: Tully Environmental, Inc

Contract: TULL01

Lab code: ACE

Sdg no.: Q3760

Instrument id number: _____

Method: _____

Run number: LB138187


Start date: 12/10/2025


End date: 12/10/2025

Lab sample id.	Client Sample Id	d/f	Time	Parameter list
S0	S0	1	1205	Cu,Fe,Pb
S1	S1	1	1209	Cu,Fe,Pb
S2	S2	1	1214	Cu,Fe,Pb
S3	S3	1	1218	Cu,Fe,Pb
S4	S4	1	1222	Cu,Fe,Pb
S5	S5	1	1226	Cu,Fe,Pb
ICV01	ICV01	1	1306	Cu,Fe,Pb
LLICV01	LLICV01	1	1316	Cu,Fe,Pb
ICB01	ICB01	1	1321	Cu,Fe,Pb
CRI01	CRI01	1	1333	Cu,Fe,Pb
ICSA01	ICSA01	1	1338	Cu,Fe,Pb
ICSAB01	ICSAB01	1	1342	Cu,Fe,Pb
ICSA	ICSA	20	1346	Cu,Fe,Pb
ICSAB	ICSAB	20	1350	Cu,Fe,Pb
CCV01	CCV01	1	1355	Cu,Fe,Pb
CCB01	CCB01	1	1359	Cu,Fe,Pb
CCV02	CCV02	1	1445	Cu,Fe,Pb
CCB02	CCB02	1	1449	Cu,Fe,Pb
Q3760-01	001 Willets Pt Blvd (Dec)	1	1520	Cu,Fe,Pb
Q3760-02	002 35th Ave (Dec)	1	1524	Cu,Fe,Pb
PB170816BL	PB170816BL	1	1529	Cu,Fe,Pb
PB170816BS	PB170816BS	1	1533	Cu,Fe,Pb
CCV03	CCV03	1	1546	Cu,Fe,Pb
CCB03	CCB03	1	1550	Cu,Fe,Pb
Q3771-02DUP	CompDUP	1	1624	Cu,Fe,Pb
Q3771-02L	CompL	5	1629	Cu,Fe,Pb
Q3771-02MS	CompMS	1	1633	Cu,Fe,Pb
CCV04	CCV04	1	1642	Cu,Fe,Pb
CCB04	CCB04	1	1646	Cu,Fe,Pb
Q3771-02MSD	CompMSD	1	1651	Cu,Fe,Pb
CCV05	CCV05	1	1749	Cu,Fe,Pb
CCB05	CCB05	1	1753	Cu,Fe,Pb
CCV06	CCV06	1	1841	Cu,Fe,Pb
CCB06	CCB06	1	1845	Cu,Fe,Pb
CCV07	CCV07	1	1933	Cu,Fe,Pb
CCB07	CCB07	1	1937	Cu,Fe,Pb
CCV08	CCV08	1	2021	Cu,Fe,Pb
CCB08	CCB08	1	2025	Cu,Fe,Pb



METAL RAW DATA






TI 190.856 {477}


Date of Fit:	12/10/2025 14:14:13	Type of Fit:	Linear	Weighting:	1/Conc
--------------	---------------------	--------------	--------	------------	--------


A0 (Offset):	0.000664	Re-Slope:	1.000000
A1 (Gain):	0.048480	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999926	Status:	OK.
Std Error of Est:	0.000019		
Predicted MDL:	0.004652		
Predicted MQL:	0.015506		


Status: OK.


Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)/IR	Std Dev	Emphasis
S0	.00000	.00000	.000	.000	.00066	.000	1
S1	.04000	.03734	-.003	-6.65	.00235	.000	1
S3	2.5000	2.4969	-.003	-.125	.11950	.002	1
S4	5.0000	4.9150	-.085	-1.70	.23452	.012	1
S5	10.000	10.097	.097	.973	.48134	.007	1
S2	.80000	.79338	-.007	-.828	.03842	.001	1





















Ca 373.690 { 90}


Date of Fit:	12/10/2025 14:14:13	Type of Fit:	Linear	Weighting:	1/Conc
--------------	---------------------	--------------	--------	------------	--------


A0 (Offset):	0.000972	Re-Slope:	1.000000
A1 (Gain):	0.056038	Y-int:	0.000000
A2 (Curvature):	0.000000		
n (Exponent):	1.000000		
Correlation:	0.999733	Status:	OK.
Std Error of Est:	0.000685		
Predicted MDL:	0.013575		
Predicted MQL:	0.045251		


Status: OK.


Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)/IR	Std Dev	Emphasis
S0	.00000	-.00023	-.000	.000	.00096	.000	1
S2	4.0000	4.2783	.278	6.96	.24072	.001	1
S3	12.500	12.985	.485	3.88	.72865	.001	1
S4	25.000	24.634	-.366	-1.46	1.3814	.004	1
S5	50.000	49.538	-.462	-.925	2.7769	.010	1
S1	2.0000	2.0646	.065	3.23	.11667	.001	1




































































Date of Fit:	12/10/2025 14:14:13	Type of Fit:	Linear	Weighting:	1/Conc
A0 (Offset):	0.000000	Re-Slope:	1.000000		
A1 (Gain):	0.000000	Y-int:	0.000000		
A2 (Curvature):	0.000000				
n (Exponent):	1.000000				
Correlation:	0.000000	Status:	Warning	Zero Gain	
Std Error of Est:	0.000000				
Predicted MDL:	n/a				
Predicted MQL:	n/a				

Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
-----------	--------------	-------------	------------	---------	-------	---------	----------




Date of Fit:	<not fit>	Type of Fit:	Linear	Weighting:	1/Conc
A0 (Offset):	0.000000	Re-Slope:	1.000000		
A1 (Gain):	0.000000	Y-int:	0.000000		





Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
-----------	--------------	-------------	------------	---------	-------	---------	----------



Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis
-----------	--------------	-------------	------------	---------	-------	---------	----------

A2 (Curvature):		0.000000		Status:	Warning	Zero Gain	
n (Exponent):		1.000000					
Correlation:		0.000000					
Std Error of Est:		0.000000					
Predicted MDL:		n/a					
Predicted MQL:		n/a					
Std. Name	Stated Conc.	Found Conc.	Difference	% Diff.	(S)IR	Std Dev	Emphasis

Water Metals Preparation Sheet

PB1708

SOP ID : M200.7-Trace Elements-22

SDG No : N/A

Matrix : WATER

Pipette ID : ICP A

Balance ID : N/A

Filter paper ID : N/A

pH Strip ID : M6069

Hood ID : #3

Block ID : 1. HOT BLOCK #1

2. N/A

Start Digest Date: 12/04/2025 Time : 10:20 Temp : 96 °C

End Digest Date: 12/04/2025 Time : 13:25 Temp : 96 °C

Digestion tube ID: M5595

Block thermometer ID: MET-DIG. #1

Dig Technician Signature: *SKS*

Supervisor Signature: *[Signature]*

Temp :

1. 96°C 2. N/A

Standard Name	MLS USED	STD REF. # FROM LOG
LFS-1	0.25	M6180
LFS-2	0.25	M6181
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Conc. HNO3	3.00	M6187
1:1 HCL	5.00	MP87148
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

HOT BLOCK#1 CELL#50 96C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
12/04/25 14:32	<i>SKS met dig</i>	<i>[Signature]</i>
	Preparation Group	Analysis Group

Lab Sample ID	Client Sample ID	pH	Initial Vol (ml)	Final Vol (ml)	Color Before	Color After	Clarity Before	Clarity After	Comment	Prep Pos
PB170816BL	PBW816	<2	50	25	Colorless	Colorless	Clear	Clear	N/A	18
PB170816BS	LCS816	<2	50	25	Colorless	Colorless	Clear	Clear	M6209,M6201	19
Q3760-01	001 WILLETS PT BLVD (DEC)	<2	50	25	Colorless	Colorless	Cloudy	Clear	N/A	20
Q3760-02	002 35TH AVE (DEC)	<2	50	25	Colorless	Colorless	Cloudy	Clear	N/A	21
Q3771-02	COMP	<2	50	25	Colorless	Colorless	Cloudy	Clear	N/A	22
Q3771-02MS	COMPMS	<2	50	25	Colorless	Colorless	Cloudy	Clear	M6209,M6201	24
Q3771-02MSD	COMPMSD	<2	50	25	Colorless	Colorless	Cloudy	Clear	M6209,M6201	25
Q3771-02DUP	COMPDUP	<2	50	25	Colorless	Colorless	Cloudy	Clear	N/A	23

WORKLIST(Hardcopy Internal Chain)

Worklist Name : PB170816

Worklist ID : 193471

Department : Digestion

Date : 12-04-2025 09:43:47

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3760-01	001 Willels Pt Blvd (Dec)	Water	Metals Group 10	1:1 HNO3 to pH < 2	TULL01	A11	12/02/2025	200.7
Q3760-02	002 35th Ave (Dec)	Water	Metals Group 10	1:1 HNO3 to pH < 2	TULL01	A11	12/02/2025	200.7
Q3771-02	Comp	Water	Metals ICP-Group1	1:1 HNO3 to pH < 2	ARAM01	A11	12/04/2025	200.7

Date/Time 12/04/2025 09:55

Raw Sample Received by: SPJmetdig

Raw Sample Relinquished by: CPCS m

Date/Time 12/04/2025 10:55

Raw Sample Received by: CPCS m

Raw Sample Relinquished by: SPJmetdig

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138187

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/10/25 12:05		Jaswal	OK
2	S1	S1	CAL2	12/10/25 12:09		Jaswal	OK
3	S2	S2	CAL3	12/10/25 12:14		Jaswal	OK
4	S3	S3	CAL4	12/10/25 12:18		Jaswal	OK
5	S4	S4	CAL5	12/10/25 12:22		Jaswal	OK
6	S5	S5	CAL6	12/10/25 12:26		Jaswal	OK
7	ICV01	ICV01	ICV	12/10/25 13:06		Jaswal	OK
8	LLICV01	LLICV01	LLICV	12/10/25 13:16		Jaswal	OK
9	ICB01	ICB01	ICB	12/10/25 13:21		Jaswal	OK
10	CRI01	CRI01	CRDL	12/10/25 13:33		Jaswal	OK
11	ICSA01	ICSA01	ICSA	12/10/25 13:38		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	12/10/25 13:42		Jaswal	OK
13	ICSADL	ICSADL	ICSA	12/10/25 13:46		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	12/10/25 13:50		Jaswal	OK
15	CCV01	CCV01	CCV	12/10/25 13:55		Jaswal	OK
16	CCB01	CCB01	CCB	12/10/25 13:59		Jaswal	OK
17	Q3763-01	MC0KC9	SAM	12/10/25 14:03		Jaswal	OK
18	Q3763-02	MC0KD0	SAM	12/10/25 14:07		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138187

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	Q3763-04	MC0KD2	SAM	12/10/25 14:16		Jaswal	OK
20	Q3763-06	MC0KE5	SAM	12/10/25 14:20		Jaswal	OK
21	Q3763-07	MC0KE6	SAM	12/10/25 14:24		Jaswal	OK
22	Q3763-11	MC0KF1	SAM	12/10/25 14:29		Jaswal	OK
23	CCV02	CCV02	CCV	12/10/25 14:45		Jaswal	OK
24	CCB02	CCB02	CCB	12/10/25 14:49		Jaswal	OK
25	Q3766-12	MC0KH8	SAM	12/10/25 14:54		Jaswal	OK
26	Q3766-02	MC0KH0	SAM	12/10/25 14:58		Jaswal	OK
27	Q3766-14	MC0KJ0	SAM	12/10/25 15:06		Jaswal	OK
28	Q3760-01	001 Willets Pt Blvd (D	SAM	12/10/25 15:20		Jaswal	OK
29	Q3760-02	002 35th Ave (Dec)	SAM	12/10/25 15:24		Jaswal	OK
30	PB170816BL	PB170816BL	MB	12/10/25 15:29		Jaswal	OK
31	PB170816BS	PB170816BS	LCS	12/10/25 15:33		Jaswal	OK
32	PB170811TB	PB170811TB	MB	12/10/25 15:37		Jaswal	OK
33	Q3770-01	4207	SAM	12/10/25 15:41		Jaswal	OK
34	CCV03	CCV03	CCV	12/10/25 15:46		Jaswal	OK
35	CCB03	CCB03	CCB	12/10/25 15:50		Jaswal	OK
36	Q3770-01DUP	4207DUP	DUP	12/10/25 15:59		Jaswal	OK
37	Q3770-01L	4207L	SD	12/10/25 16:03		Jaswal	OK
38	Q3770-01MS	4207MS	MS	12/10/25 16:07		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138187

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	Q3770-01MSD	4207MSD	MSD	12/10/25 16:12		Jaswal	OK
40	Q3770-01A	4207A	PS	12/10/25 16:16		Jaswal	OK
41	Q3771-02	Comp	SAM	12/10/25 16:20		Jaswal	OK
42	Q3771-02DUP	CompDUP	DUP	12/10/25 16:24		Jaswal	OK
43	Q3771-02L	CompL	SD	12/10/25 16:29		Jaswal	OK
44	Q3771-02MS	CompMS	MS	12/10/25 16:33		Jaswal	OK
45	CCV04	CCV04	CCV	12/10/25 16:42		Jaswal	OK
46	CCB04	CCB04	CCB	12/10/25 16:46		Jaswal	OK
47	Q3771-02MSD	CompMSD	MSD	12/10/25 16:51		Jaswal	OK
48	Q3771-02A	CompA	PS	12/10/25 16:55		Jaswal	OK
49	Q3763-05	MC0KD7	SAM	12/10/25 16:59		Jaswal	OK
50	Q3766-06	MC0KH4	SAM	12/10/25 17:03		Jaswal	OK
51	Q3766-09	MC0KH7	SAM	12/10/25 17:08		Jaswal	OK
52	LR-1	LR-1	HIGH STD	12/10/25 17:12		Jaswal	OK
53	LR-2	LR-2	HIGH STD	12/10/25 17:26		Jaswal	OK
54	Q3766-03	MC0KH1	SAM	12/10/25 17:35		Jaswal	OK
55	LR-3	LR-3	HIGH STD	12/10/25 17:40		Jaswal	OK
56	Q3795-04	72-11966	SAM	12/10/25 17:44		Jaswal	OK
57	CCV05	CCV05	CCV	12/10/25 17:49		Jaswal	OK
58	CCB05	CCB05	CCB	12/10/25 17:53		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138187

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

59	Q3795-08	RBR-200059	SAM	12/10/25 17:57		Jaswal	OK
60	Q3803-01	CHASE-A	SAM	12/10/25 18:02		Jaswal	OK
61	Q3803-02	CHASE-B	SAM	12/10/25 18:06		Jaswal	OK
62	Q3803-03	CHASE-C	SAM	12/10/25 18:10		Jaswal	OK
63	Q3803-04	CHASE-D	SAM	12/10/25 18:15		Jaswal	OK
64	Q3803-05	CHASE-E	SAM	12/10/25 18:19		Jaswal	OK
65	Q3803-06	CHASE-F	SAM	12/10/25 18:23		Jaswal	OK
66	Q3803-07	CHASE-G	SAM	12/10/25 18:28		Jaswal	OK
67	Q3803-08	CHASE-H	SAM	12/10/25 18:32		Jaswal	OK
68	Q3803-09	CHASE-I	SAM	12/10/25 18:37		Jaswal	OK
69	CCV06	CCV06	CCV	12/10/25 18:41		Jaswal	OK
70	CCB06	CCB06	CCB	12/10/25 18:45		Jaswal	OK
71	Q3803-10	CHASE-J	SAM	12/10/25 18:49		Jaswal	OK
72	Q3803-11	CHASE-K	SAM	12/10/25 18:54		Jaswal	OK
73	Q3803-12	CHASE-L	SAM	12/10/25 18:58		Jaswal	OK
74	Q3803-13	CHASE-M	SAM	12/10/25 19:02		Jaswal	OK
75	Q3803-14	CHASE-N	SAM	12/10/25 19:07		Jaswal	OK
76	Q3810-01	BP Dirt Sample	SAM	12/10/25 19:11		Jaswal	OK
77	Q3810-01DUP	BP Dirt SampleDUP	DUP	12/10/25 19:16		Jaswal	OK
78	Q3810-01L	BP Dirt SampleL	SD	12/10/25 19:20		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB138187

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

79	Q3810-01MS	BP Dirt SampleMS	MS	12/10/25 19:24		Jaswal	OK
80	Q3810-01MSD	BP Dirt SampleMSD	MSD	12/10/25 19:28		Jaswal	OK
81	CCV07	CCV07	CCV	12/10/25 19:33		Jaswal	OK
82	CCB07	CCB07	CCB	12/10/25 19:37		Jaswal	OK
83	Q3810-01A	BP Dirt SampleA	PS	12/10/25 19:41		Jaswal	OK
84	Q3791-01	T2	SAM	12/10/25 19:51		Jaswal	OK
85	PB170857TB	PB170857TB	MB	12/10/25 19:55		Jaswal	OK
86	Q3790-02	304 FURMAN SOIL	SAM	12/10/25 20:00		Jaswal	OK
87	PB170887BL	PB170887BL	MB	12/10/25 20:04		Jaswal	OK
88	PB170887BS	PB170887BS	LCS	12/10/25 20:08		Jaswal	OK
89	Q3780-02	PLASTIC SHEETING	SAM	12/10/25 20:13		Jaswal	OK
90	Q3812-01	OR-02-120825	SAM	12/10/25 20:17		Jaswal	OK
91	CCV08	CCV08	CCV	12/10/25 20:21		Jaswal	OK
92	CCB08	CCB08	CCB	12/10/25 20:25		Jaswal	OK



SHIPPING DOCUMENTS

CLIENT INFORMATION

REPORT TO BE SENT TO:

COMPANY: Tully Env. Remediation Inc
ADDRESS: 127-50 Northern Blvd
CITY: Flushing STATE: NY ZIP: 11358
ATTENTION: D Devo
PHONE: 718 446 7000 FAX:

CLIENT PROJECT INFORMATION

PROJECT NAME: Transfer Station SPDES
PROJECT NO.: 252113 LOCATION:
PROJECT MANAGER:
e-mail:
PHONE: FAX:

CLIENT BILLING INFORMATION

BILL TO: Same PO#:
ADDRESS:
CITY STATE: ZIP:
ATTENTION: PHONE:

ANALYSIS

DATA TURNAROUND INFORMATION

FAX (RUSH) _____ DAYS*
HARDCOPY (DATA PACKAGE): _____ DAYS*
EDD: _____ DAYS*

*TO BE APPROVED BY CHEMTECH
STANDARD HARDCOPY TURNAROUND TIME IS 10 BUSINESS

DATA DELIVERABLE INFORMATION

☒ Level 1 (Results Only) ☐ Level 4 (QC + Full Raw Data)
☐ Level 2 (Results + QC) ☐ NJ Reduced ☐ US EPA CLP
☐ Level 3 (Results + QC) ☐ NYS ASP A ☐ NYS ASP B
+ Raw Data ☐ Other _____
☐ EDD FORMAT

1: 2: 3: 4: 5: 6: 7: 8: 9:
Ammonia DO Cu-Fe-Pb BTEX Hg 1631LL BUDDS

PRESERVATIVES

COMMENTS

ALLIANCE SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE		SAMPLE COLLECTION		# OF BOTTLES	PRESERVATIVES									COMMENTS	
			COMP	GRAB	DATE	TIME		1	2	3	4	5	6	7	8	9	← Specify Preservatives A-HCl D-NaOH B-HNO3 E-ICE C-H2SO4 F-OTHER	
1.	001 Willets Pt Blvd (Dec)	W		X	12/2	130		X	X	X	X	X	X					
2.	002 35th Ave (Dec)	W		X	12/2	130		X	X	X	X	X	X					
3.																		
4.																		
5.																		
6.																		
7.																		
8.																		
9.																		
10.																		

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE POSSESSION INCLUDING COURIER DELIVERY

RELINQUISHED BY SAMPLER: 1. <u>D Devo</u>	DATE/TIME: <u>12/2/25</u>	RECEIVED BY: 1. _____	Conditions of bottles or coolers at receipt: ☐ COMPLIANT ☐ NON COMPLIANT ☐ COOLER TEMP <u>5.3</u> °C
RELINQUISHED BY SAMPLER: 2. _____	DATE/TIME: <u>12/3/25 11:11</u>	RECEIVED BY: 2. <u>CD</u>	Comments: <u>IF Env #1</u>
RELINQUISHED BY SAMPLER: 3. _____	DATE/TIME: _____	RECEIVED BY: 3. _____	Page ____ of CLIENT: ☐ Hand Delivered ☐ Other Shipment Complete ☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q3760 TULL01
Client Name : Tully Environmental, Inc
Client Contact : Dean Devoc
Invoice Name : Tully Environmental, Inc
Invoice Contact : Dean Devoc

Order Date : 12/3/2025 10:53:00 AM
Project Name : Transfer Station-SPDES
Receive DateTime : 12/3/2025 2:00:00 PM
Purchase Order : 11:15 AM

Project Mgr :
Report Type : Results Only
EDD Type : EXCEL NOCLEANUP
Hard Copy Date :
Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q3760-01	001 Willets Pt Blvd (Dec) <i>CDCC</i>	Water	12/02/2025	13:00	VOC-BTEX		624.1	5 Bus. Days	
Q3760-02	002 35th Ave (Dec)	Water	12/02/2025	13:00	VOC-BTEX		624.1	5 Bus. Days	

Relinquished By : *CD*

Date / Time : 12/3/25 1320

Received By : *SPM*

Date / Time : 12/3/25 1352

Storage Area : VOA Refridgerator Room