

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Copper	986	1000	99	95 - 105	P	12/10/2025	13:06	LB138187
	Iron	3840	4000	96	95 - 105	P	12/10/2025	13:06	LB138187
	Lead	4030	4000	101	95 - 105	P	12/10/2025	13:06	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Copper	21.5	20.0	107	80 - 120	P	12/10/2025	13:16	LB138187
	Iron	105	100	105	80 - 120	P	12/10/2025	13:16	LB138187
	Lead	12.2	12.0	102	80 - 120	P	12/10/2025	13:16	LB138187

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: Tully Environmental, Inc

SDG No.: Q3760

Contract: TULL01

Lab Code: ACE

Initial Calibration Source: EPA

Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Copper	1250	1250	100	90 - 110	P	12/10/2025	13:55	LB138187
	Iron	5080	5000	102	90 - 110	P	12/10/2025	13:55	LB138187
	Lead	4970	5000	99	90 - 110	P	12/10/2025	13:55	LB138187
CCV02	Copper	1250	1250	100	90 - 110	P	12/10/2025	14:45	LB138187
	Iron	5130	5000	103	90 - 110	P	12/10/2025	14:45	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	14:45	LB138187
CCV03	Copper	1260	1250	101	90 - 110	P	12/10/2025	15:46	LB138187
	Iron	5040	5000	101	90 - 110	P	12/10/2025	15:46	LB138187
	Lead	4930	5000	99	90 - 110	P	12/10/2025	15:46	LB138187
CCV04	Copper	1240	1250	99	90 - 110	P	12/10/2025	16:42	LB138187
	Iron	4920	5000	98	90 - 110	P	12/10/2025	16:42	LB138187
	Lead	4890	5000	98	90 - 110	P	12/10/2025	16:42	LB138187
CCV05	Copper	1240	1250	99	90 - 110	P	12/10/2025	17:49	LB138187
	Iron	5010	5000	100	90 - 110	P	12/10/2025	17:49	LB138187
	Lead	4860	5000	97	90 - 110	P	12/10/2025	17:49	LB138187
CCV06	Copper	1220	1250	98	90 - 110	P	12/10/2025	18:41	LB138187
	Iron	4780	5000	96	90 - 110	P	12/10/2025	18:41	LB138187
	Lead	4820	5000	96	90 - 110	P	12/10/2025	18:41	LB138187
CCV07	Copper	1220	1250	98	90 - 110	P	12/10/2025	19:33	LB138187
	Iron	4800	5000	96	90 - 110	P	12/10/2025	19:33	LB138187
	Lead	4790	5000	96	90 - 110	P	12/10/2025	19:33	LB138187
CCV08	Copper	1240	1250	99	90 - 110	P	12/10/2025	20:21	LB138187
	Iron	4860	5000	97	90 - 110	P	12/10/2025	20:21	LB138187
	Lead	4870	5000	98	90 - 110	P	12/10/2025	20:21	LB138187